



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB117662

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	12/06/2021	12:40
2	M4916-01		1	10	100.00	>212.0	>212.0	12/06/2021	13:05
3	M4916-01DUP		1	9	100.00	>212.0	>212.0	12/06/2021	13:35
4	M4934-03		1	8	40.00	104.0	103.7	12/06/2021	14:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB117662

WorkList Name : fp12621-1 WorkList ID : 154899 Department : Wet-Chemistry Date : 12-06-2021 09:29:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/16/2021	Solid	M4916-01	Flash Point	Cool 4 deg C	PSEG03	F41	BUR-9890	12/02/2021	1010B
12/10/2021	Solid	M4934-03	Flash Point	Cool 4 deg C	PSEG03	F42	17077	12/03/2021	1010B

Date/Time 12/06/21 12:15
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: RJ (DFA-166)

Date/Time 12/06/21 14:15
 Raw Sample Received by: RJ (DFA-196)
 Raw Sample Relinquished by: KS (WC)