



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB117663

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	12/06/2021	12:40
2	M4920-04		1	12	100.00	>212.0	>212.0	12/06/2021	14:40
3	M4920-04DUP		1	12	100.00	>212.0	>212.0	12/06/2021	15:05
4	M4934-05		1	15	100.00	>212.0	>212.0	12/06/2021	15:35
5	M4934-06		1	14	100.00	>212.0	>212.0	12/06/2021	16:10
6	M4934-07		1	13	100.00	>212.0	>212.0	12/06/2021	16:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB117663

WorkList Name : fp12621 WorkList ID : 154898 Department : Wet-Chemistry Date : 12-06-2021 09:28:49

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/09/2021	Water	M4920-04	Flash Point	Cool 4 deg C	PARS02	D11	FDR-20211202-WC-W	12/02/2021	1010B
12/10/2021	Water	M4934-05	Flash Point	Cool 4 deg C	PSEG03	F42	17078	12/03/2021	1010B
12/10/2021	Water	M4934-06	Flash Point	Cool 4 deg C	PSEG03	F42	17094	12/03/2021	1010B
12/10/2021	Water	M4934-07	Flash Point	Cool 4 deg C	PSEG03	F42	17097	12/03/2021	1010B

Date/Time: 12/06/21 14:15
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: RJ (KST-106)

Date/Time: 12/06/21 16:45
 Raw Sample Received by: RJ (KST-106)
 Raw Sample Relinquished by: KS (WC)