



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117666

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/06/2021	16:45
2	CAL2	1	Water	NA	NA	20.3	7.00	12/06/2021	16:46
3	CAL3	1	Water	NA	NA	20.4	10.03	12/06/2021	16:47
4	ICV	1	Water	NA	NA	20.5	7.02	12/06/2021	16:50
5	CCV1	1	Water	NA	NA	20.3	2.01	12/06/2021	16:52
6	M4953-01	1	Water	NA	NA	20.1	10.72	12/06/2021	17:00
7	M4953-01DUP	1	Water	NA	NA	20.2	10.73	12/06/2021	17:01
8	M4953-02	1	Water	NA	NA	20.3	11.50	12/06/2021	17:05
9	M4953-03	1	Water	NA	NA	20.3	11.35	12/06/2021	17:10
10	CCV2	1	Water	NA	NA	20.5	12.03	12/06/2021	17:11

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M4953 WorkList ID : 154906 Department : Wet-Chemistry Date : 12-06-2021 16:28:17

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/06/2021	Water	M4953-01	OK pH	Cool 4 deg C	PSEG03	F52	A377	12/06/2021	9040C
12/06/2021	Water	M4953-02	OK pH	Cool 4 deg C	PSEG03	F52	A3718	12/06/2021	9040C
12/06/2021	Water	M4953-03	OK pH	Cool 4 deg C	PSEG03	F52	A1513	12/06/2021	9040C

Date/Time 12/06/21 16:30
 Raw Sample Received by: JR WCC
 Raw Sample Relinquished by: CP SM

Date/Time 12/06/21 17:45
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JR WCC