



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: jigneshDate: 12/07/2021OVEN IN TEMP1: 104 °COVEN OUT TEMP1: 103 °CDATE & TIME IN1: 12/06/2021 17:00 DATE & TIME OUT1: 12/07/2021 08:00OVEN IN TEMP2: 104 °COVEN OUT TEMP2: 104 °CDATE & TIME IN2: 12/07/2021 08:02 DATE & TIME OUT2: 12/07/2021 10:00Run Number: LB117667BalanceID: WC SC-6OvenID: WC OVEN-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB117667BL	LB117667BL	95.3147	95.3147	100	95.3148	95.3148	95.3148	0.0001	1
2	M4953-01	A377	89.1482	89.1482	100	89.2043	89.2043	89.2043	0.0561	561
3	M4953-02	A3718	75.5422	75.5422	100	75.6150	75.6150	75.6150	0.0728	728
4	M4953-03	A1513	85.0749	85.0749	100	85.1547	85.1547	85.1547	0.0798	798
5	M4953-03DUP	A1513DUP	81.4770	81.4770	100	81.5568	81.5568	81.5568	0.0798	798

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

WORKLIST(Hardcopy Internal Chain)

2P 11766F

WorkList Name : TS M4953 W

WorkList ID : 154907

Department : Wet-Chemistry

Date : 12-06-2021 16:28:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/09/2021	Water	M4953-01	TS	Cool 4 deg C	PSEG03	F52	A377	12/06/2021	SM2540 B
12/09/2021	Water	M4953-02	TS	Cool 4 deg C	PSEG03	F52	A3718	12/06/2021	SM2540 B
12/09/2021	Water	M4953-03	TS	Cool 4 deg C	PSEG03	F52	A1513	12/06/2021	SM2540 B

Date/Time 12/06/21 16:30
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: CP SM

Date/Time 12/06/21 18:00
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JP (WC)