

Analytical Summary ReportAnalysis Method: SM5220 DANALYST: IwonaParameter: CODSUPERVISOR REVIEW BY: apatelRun Number: LB117684

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP93829
COD calibration std. 100 ppm	WP93827
COD calibration std. 10 ppm	WP93825
COD calibration std. 150 ppm	WP93828
COD calibration std. 50 ppm	WP93826
COD calibration std. 0 ppm	WP93824
COD CCV std, 50ppm	WP93830
COD Digestion Vials Low Level 0-150Mg/L	W2826

Temp In (C): <u>148</u>	Date In: <u>12/07/2021</u>	Time In: <u>10:30</u>
Temp Out (C): <u>150</u>	Date Out: <u>12/07/2021</u>	Time Out: <u>12:30</u>

Intercept: -0.2481Slope: 1.0105Regression: 1

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	0.246		11/09/2021	13:10
2	CAL2	10	1	Water	9.000	9.152	8.5	11/09/2021	13:10
3	CAL3	50	1	Water	51.000	50.716	-1.4	11/09/2021	13:11
4	CAL4	100	1	Water	101.000	100.196	-0.2	11/09/2021	13:11
5	CAL5	150	1	Water	151.000	149.676	0.2	11/09/2021	13:12



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Iwona

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB117684

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50	NA	NA	1	Water	50.000	49.726	11/09/2021	13:16
2	ICB		NA	NA	1	Water	0.000	0.246	11/09/2021	13:16
3	CCV1	50	NA	NA	1	Water	50.000	49.726	12/08/2021	10:05
4	CCB1		NA	NA	1	Water	0.000	0.246	12/08/2021	10:05
5	LB117684BL		NA	NA	1	Water	0.000	0.246	12/08/2021	10:06
6	LB117684BS	50	NA	NA	1	Water	50.000	49.726	12/08/2021	10:06
7	M4953-01		NA	NA	1	Water	80.000	79.414	12/08/2021	10:07
8	M4953-01DUP		NA	NA	1	Water	81.000	80.404	12/08/2021	10:07
9	M4953-01MS	50	NA	NA	1	Water	125.000	123.947	12/08/2021	10:08
10	M4953-01MSD	50	NA	NA	1	Water	126.000	124.936	12/08/2021	10:08
11	M4953-02		NA	NA	1	Water	136.000	134.832	12/08/2021	10:09
12	M4953-03		NA	NA	1	Water	142.000	140.770	12/08/2021	10:09
13	CCV2	50	NA	NA	1	Water	50.000	49.726	12/08/2021	10:10
14	CCB2		NA	NA	1	Water	0.000	0.246	12/08/2021	10:10