


LB117696

7470A

INSTRUMENT ID: CV2

Linear



A= 0.0000e+000

B= 2.3880e-004

C= 2.6227e-002

Rho= 0.9999605

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.037	0.037	45	0.000	45					1
0.2	0.200	0.201	0.001	731	0.0 %	731					0
2.5	2.500	2.493	-0.007	10331	0.0 %	10331					0
5.0	5.000	4.941	-0.059	20581	0.0 %	20581					1
7.5	7.500	7.490	-0.010	31254	0.0 %	31254					0
10.0	10.000	10.038	0.038	41926	0.0 %	41926					0

LB117696 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 08 Dec 2021 11:33:51

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	45	-	0.0000	7470A	PPB	08 Dec 2021 11:40:19	S	Std
0.2 - 1	50.2	731	-	0.2000	7470A	PPB	08 Dec 2021 11:42:35	S	Std
2.5 - 1	52.5	10331	-	2.5000	7470A	PPB	08 Dec 2021 11:44:53	S	Std
5.0 - 1	55	20581	-	5.0000	7470A	PPB	08 Dec 2021 11:47:11	S	Std
7.5 - 1	57.5	31254	-	7.5000	7470A	PPB	08 Dec 2021 11:49:27	S	Std
10.0 - 1	510	41926	-	10.0000	7470A	PPB	08 Dec 2021 11:51:43	S	Std
ICV01 - 1	ICV01	15213	3.6591	-	7470A	PPB	08 Dec 2021 11:54:24	U	SMPL
ICB01 - 1	ICB01	-32	0.0186	-	7470A	PPB	08 Dec 2021 11:56:39	U	SMPL
CCV75 - 1	CCV75	19490	4.6805	-	7470A	PPB	08 Dec 2021 11:58:54	U	SMPL
CCB75 - 1	CCB75	-24	0.0205	-	7470A	PPB	08 Dec 2021 12:01:09	U	SMPL
CRA - 1	CRA	686	0.1900	-	7470A	PPB	08 Dec 2021 12:03:24	U	SMPL
HighStd - 1	HighStd	39507	9.4606	-	7470A	PPB	08 Dec 2021 12:05:38	U	SMPL
ChkStd - 1	ChkStd	28453	6.8208	-	7470A	PPB	08 Dec 2021 12:07:53	U	SMPL
PB141235BL - 1	PBW	-49	0.0145	-	7470A	PPB	08 Dec 2021 12:10:09	U	SMPL
PB141235BS - 1	LCSW	15885	3.8196	-	7470A	PPB	08 Dec 2021 12:12:25	U	SMPL
M4953-01 - 1	A377	156	0.0635	-	7470A	PPB	08 Dec 2021 12:14:42	U	SMPL
M4953-01DUP - 1	A377DUP	76	0.0444	-	7470A	PPB	08 Dec 2021 12:16:58	U	SMPL
M4953-01MS - 1	A377MS	13855	3.3348	-	7470A	PPB	08 Dec 2021 12:19:15	U	SMPL
M4953-01MSD - 1	A377MSD	14462	3.4798	-	7470A	PPB	08 Dec 2021 12:21:32	U	SMPL
M4953-02 - 1	A3718	-2	0.0257	-	7470A	PPB	08 Dec 2021 12:23:48	U	SMPL
CCV76 - 1	CCV76	19971	4.7953	-	7470A	PPB	08 Dec 2021 12:26:03	U	SMPL
CCB76 - 1	CCB76	-26	0.0200	-	7470A	PPB	08 Dec 2021 12:28:18	U	SMPL
M4953-03 - 1	A1513	82	0.0458	-	7470A	PPB	08 Dec 2021 12:30:34	U	SMPL
M4953-01LX5 - 1		32	0.0339	-	7470A	PPB	08 Dec 2021 12:35:25	U	SMPL
M4953-01A - 1		16304	3.9196	-	7470A	PPB	08 Dec 2021 12:37:40	U	SMPL
CCV77 - 1	CCV77	20164	4.8414	-	7470A	PPB	08 Dec 2021 12:39:55	U	SMPL
CCB77 - 1	CCB77	-32	0.0186	-	7470A	PPB	08 Dec 2021 12:42:11	U	SMPL