



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB117708

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	12/07/2021	14:05
2	M4953-01		1	10	100.00	>212.0	>212.0	12/07/2021	14:35
3	M4953-01DUP		1	10	100.00	>212.0	>212.0	12/07/2021	15:10
4	M4953-02		1	12	100.00	>212.0	>212.0	12/07/2021	15:35
5	M4953-03		1	15	100.00	>212.0	>212.0	12/07/2021	15:58

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP 12721 WorkList ID : 154934 Department : Wet-Chemistry Date : 12-07-2021 13:40:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/09/2021	Water	M4953-01	Flash Point	Cool 4 deg C	PSEG03	F52	A377	12/06/2021	1010B
12/09/2021	Water	M4953-02	Flash Point	Cool 4 deg C	PSEG03	F52	A3718	12/06/2021	1010B
12/09/2021	Water	M4953-03	Flash Point	Cool 4 deg C	PSEG03	F52	A1513	12/06/2021	1010B

Date/Time 12/07/21 13:45
 Raw Sample Received by: KS(WC)
 Raw Sample Relinquished by: RJCEA-1061

Date/Time 12/07/21 16:05
 Raw Sample Received by: RJCEA-1061
 Raw Sample Relinquished by: KS(WC)