



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB117709

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/08/2021	15:40
2	CAL2	1	Water	NA	NA	20.2	7.00	12/08/2021	15:41
3	CAL3	1	Water	NA	NA	20.2	10.03	12/08/2021	15:42
4	ICV	1	Water	NA	NA	20.3	7.02	12/08/2021	15:45
5	CCV1	1	Water	NA	NA	20.4	2.01	12/08/2021	15:46
6	M4966-06	1	Solid	20.03	20	20.6	5.56	12/08/2021	15:49
7	M4966-06DUP	1	Solid	20.04	20	20.7	5.57	12/08/2021	15:50
8	M4966-15	1	Solid	20.02	20	20.1	5.58	12/08/2021	16:00
9	M4966-23	1	Solid	20.03	20	20.3	7.82	12/08/2021	16:01
10	M4966-30	1	Solid	20.05	20	20.3	7.55	12/08/2021	16:03
11	M4967-02	1	Solid	20.02	20	20.4	5.55	12/08/2021	16:10
12	M4967-05	1	Solid	20.03	20	20.7	5.56	12/08/2021	16:11
13	M4967-08	1	Solid	20.05	20	20.8	5.65	12/08/2021	16:15
14	M4967-11	1	Solid	20.04	20	20.8	5.54	12/08/2021	16:20
15	M4967-14	1	Solid	20.02	20	20.3	5.58	12/08/2021	16:22
16	CCV2	1	Water	NA	NA	20.5	12.03	12/08/2021	16:25

WORKLIST(Hardcopy Internal Chain)

117709

WorkList Name : CORROSIONITY M4966

WorkList ID : 154971

Department : Wet-Chemistry

Date : 12-08-2021 13:25:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/07/2021	Solid	M4966-06	Corrosivity	Cool 4 deg C	BROW03	H11	P1-COMP	12/07/2021	9045D
12/07/2021	Solid	M4966-15	Corrosivity	Cool 4 deg C	BROW03	H11	P2-COMP	12/07/2021	9045D
12/07/2021	Solid	M4966-23	Corrosivity	Cool 4 deg C	BROW03	H11	P3-COMP	12/07/2021	9045D
12/07/2021	Solid	M4966-30	Corrosivity	Cool 4 deg C	BROW03	H11	P4-COMP	12/07/2021	9045D
12/06/2021	Solid	M4967-02	Corrosivity	Cool 4 deg C	PSEG03	H11	TP-7S	12/06/2021	9045D
12/06/2021	Solid	M4967-05	Corrosivity	Cool 4 deg C	PSEG03	H11	TP-7D	12/06/2021	9045D
12/06/2021	Solid	M4967-08	Corrosivity	Cool 4 deg C	PSEG03	H11	TP-8S	12/06/2021	9045D
12/06/2021	Solid	M4967-11	Corrosivity	Cool 4 deg C	PSEG03	H11	TP-11S	12/06/2021	9045D
12/06/2021	Solid	M4967-14	Corrosivity	Cool 4 deg C	PSEG03	H11	TP-11D	12/06/2021	9045D

Date/Time 12/08/21 13:25
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: CP SM

Date/Time 12/08/21 17:00
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JP (WC)