


LB117712

245.1 INSTRUMENT ID: CVJ

Linear



A= 0.0000e+000

B= 2.9162e-004

C= -3.6922e-002

Rho= 0.9997730

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	O/D
0.0	0.000	-0.011	-0.011	88	0.000	88					-8
0.05	0.050	0.046	-0.004	286	0.0 %	286	0				-8
0.2	0.200	0.184	-0.016	757	0.0 %	757					0
2.5	2.500	2.507	0.007	8722	0.0 %	8722					0
5.0	5.000	4.976	-0.024	17190	0.0 %	17190					0
7.5	7.500	7.668	0.168	26423	0.0 %	26423					2
10.0	10.000	9.880	-0.120	34007	0.0 %	34007					-1

MB
12/9/21

LB117712 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 09 Dec 2021 12:23:08

Sample ID	Extended ID	µ Abs.	Conc.	Strnd CondMethod	Units	Date	Type	Type
0.0 - 1	50	88	-	0.0000245.1	PPB	09 Dec 2021 13:26:13	S	Std
0.05 - 1	50.05	286	-	0.0500245.1	PPB	09 Dec 2021 13:28:30	S	Std
0.2 - 1	50.2	757	-	0.2000245.1	PPB	09 Dec 2021 13:30:47	S	Std
2.5 - 1	52.5	8722	-	2.5000245.1	PPB	09 Dec 2021 13:33:04	S	Std
5.0 - 1	55	17190	-	5.0000245.1	PPB	09 Dec 2021 13:35:20	S	Std
7.5 - 1	57.5	26423	-	7.5000245.1	PPB	09 Dec 2021 13:37:36	S	Std
10.0 - 1	510	34007	-	10.0000245.1	PPB	09 Dec 2021 13:42:30	S	Std
ICV03 - 1	ICV03	13931	4.0256	-245.1	PPB	09 Dec 2021 13:45:20	U	SMPL
ICB03 - 1	ICB03	104	-0.0066	-245.1	PPB	09 Dec 2021 13:47:34	U	SMPL
CCV82 - 1	CCV82	17799	5.1535	-245.1	PPB	09 Dec 2021 13:49:49	U	SMPL
CCB82 - 1	CCB82	84	-0.0124	-245.1	PPB	09 Dec 2021 13:52:04	U	SMPL
CRA - 1	CRA	273	0.0427	-245.1	PPB	09 Dec 2021 13:54:19	U	SMPL
HighStd - 1	HighStd	34159	9.9244	-245.1	PPB	09 Dec 2021 13:56:34	U	SMPL
ChkStd - 1	ChkStd	25292	7.3386	-245.1	PPB	09 Dec 2021 13:58:49	U	SMPL
PB141285BL - 1	PBW	215	0.0258	-245.1	PPB	09 Dec 2021 14:01:07	U	SMPL
PB141285BS - 1	LCSW	13410	3.8736	-245.1	PPB	09 Dec 2021 14:03:23	U	SMPL
M4882-02 - 1	COMP	89	-0.0110	-245.1	PPB	09 Dec 2021 14:05:39	U	SMPL
M4882-02DUP - 1	COMPDUP	89	-0.0110	-245.1	PPB	09 Dec 2021 14:07:56	U	SMPL
M4882-02MS - 1	COMPMS	13423	3.8774	-245.1	PPB	09 Dec 2021 14:13:30	U	SMPL
M4882-02MSD - 1	COMPMSD	12779	3.6896	-245.1	PPB	09 Dec 2021 14:15:47	U	SMPL
M4914-01 - 1	SEMI-ANNUAL-CONDENSATE	902	0.2261	-245.1	PPB	09 Dec 2021 14:18:03	U	SMPL
CCV83 - 1	CCV83	18135	5.2515	-245.1	PPB	09 Dec 2021 14:20:18	U	SMPL
CCB83 - 1	CCB83	86	-0.0118	-245.1	PPB	09 Dec 2021 14:22:34	U	SMPL
M4882-02LX5 - 1		58	-0.0200	-245.1	PPB	09 Dec 2021 14:24:49	U	SMPL
M4882-02A - 1		14342	4.1454	-245.1	PPB	09 Dec 2021 14:27:04	U	SMPL
CCV84 - 1	CCV84	17993	5.2101	-245.1	PPB	09 Dec 2021 14:29:20	U	SMPL
CCB84 - 1	CCB84	92	-0.0101	-245.1	PPB	09 Dec 2021 14:31:35	U	SMPL