

LB117721


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

54317



A= 0.0000e+000

B= 1.8194e-004

C= -2.4711e-002

Rho= 0.9996494

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	σ/σ)
0.0	0.000	-0.019	-0.019	34	0.000	34					5
0.2	0.200	0.190	-0.010	1182	0.0 %	1182					1
2.5	2.500	2.485	-0.015	13793	0.0 %	13793					0
5.0	5.000	4.995	-0.005	27589	0.0 %	27589					3
7.5	7.500	7.691	0.191	42406	0.0 %	42406					1
10.0	10.000	9.858	-0.142	54317	0.0 %	54317					

LB117721 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 09 Dec 2021 14:42:20					
Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type	
0.0 - 1	50	34	-	0.00007471B	PPB	09 Dec 2021 14:43:23	S	Std	
0.2 - 1	51.2	1182	-	0.20007471B	PPB	09 Dec 2021 14:45:40	S	Std	
2.5 - 1	52.5	13793	-	2.50007471B	PPB	09 Dec 2021 14:47:57	S	Std	
5.0 - 1	55	27589	-	5.00007471B	PPB	09 Dec 2021 14:50:14	S	Std	
7.5 - 1	57.5	42406	-	7.50007471B	PPB	09 Dec 2021 14:52:30	S	Std	
10.0 - 1	510	54317	-	10.00007471B	PPB	09 Dec 2021 14:54:46	S	Std	
ICV04 - 1	ICV04	21755	3.9334	-7471B	PPB	09 Dec 2021 14:57:38	U	SMPL	
ICB04 - 1	ICB04	-38	-0.0316	-7471B	PPB	09 Dec 2021 14:59:53	U	SMPL	
CCV85 - 1	CCV85	29037	5.2583	-7471B	PPB	09 Dec 2021 15:02:07	U	SMPL	
CCB85 - 1	CCB85	-38	-0.0316	-7471B	PPB	09 Dec 2021 15:04:23	U	SMPL	
CRA - 1	CRA	1169	0.1880	-7471B	PPB	09 Dec 2021 15:06:38	U	SMPL	
HighStd - 1	HighStd	54713	9.9299	-7471B	PPB	09 Dec 2021 15:08:53	U	SMPL	
ChkStd - 1	ChkStd	40047	7.2615	-7471B	PPB	09 Dec 2021 15:11:08	U	SMPL	
PB141283BL - 1	PBS	-88	-0.0407	-7471B	PPB	09 Dec 2021 15:13:25	U	SMPL	
PB141283BS - 1	LCSS	20056	3.6243	-7471B	PPB	09 Dec 2021 15:15:41	U	SMPL	
M4965-01 - 1	RED-PAINTCHIPS-EXTERIOR	696	0.1019	-7471B	PPB	09 Dec 2021 15:17:57	U	SMPL	
M4965-01DUP - 1	RED-PAINTCHIPS-EXTERIORDUP	677	0.0985	-7471B	PPB	09 Dec 2021 15:20:14	U	SMPL	
M4965-01MS - 1	RED-PAINTCHIPS-EXTERIORMS	20021	3.6179	-7471B	PPB	09 Dec 2021 15:22:30	U	SMPL	
M4965-02 - 1	RED-CONCRETE-COMP-EXTERIOR	470	0.0608	-7471B	PPB	09 Dec 2021 15:27:03	U	SMPL	
CCV86 - 1	CCV86	27704	5.0158	-7471B	PPB	09 Dec 2021 15:29:18	U	SMPL	
CCB86 - 1	CCB86	-45	-0.0329	-7471B	PPB	09 Dec 2021 15:31:33	U	SMPL	
M4965-07 - 1	WHITE-PAINTCHIPS-EXTERIOR	1132	0.1812	-7471B	PPB	09 Dec 2021 15:33:49	U	SMPL	
M4965-08 - 1	WHITE-CONCRETE-COMP-EXTERIOR	88	-0.0087	-7471B	PPB	09 Dec 2021 15:36:04	U	SMPL	
M4965-13 - 1	BLUE-PAINTCHIPS	23470	4.2455	-7471B	PPB	09 Dec 2021 15:38:20	U	SMPL	
M4965-14 - 1	BLUE-CONCRETE-COMP	1131	0.1811	-7471B	PPB	09 Dec 2021 15:40:35	U	SMPL	
M4965-19 - 1	RED-WALL-PAINTCHIPS	837	0.1276	-7471B	PPB	09 Dec 2021 15:42:51	U	SMPL	
M4965-20 - 1	RED-WALL-CONCRETE-COMP	1395	0.2291	-7471B	PPB	09 Dec 2021 15:45:07	U	SMPL	
M4965-25 - 1	WHITE-BROWN-PAINTCHIPS	1106	0.1765	-7471B	PPB	09 Dec 2021 15:47:23	U	SMPL	
M4965-26 - 1	WHITE-BROWN-CONCRETE-COMP	470	0.0608	-7471B	PPB	09 Dec 2021 15:49:40	U	SMPL	
M4965-31 - 1	FLOORS-COMP	1621	0.2702	-7471B	PPB	09 Dec 2021 15:51:57	U	SMPL	
PB141284BL - 1	PBS	7	-0.0234	-7471B	PPB	09 Dec 2021 15:54:13	U	SMPL	
CCV87 - 1	CCV87	28929	5.2387	-7471B	PPB	09 Dec 2021 16:10:20	U	SMPL	
CCB87 - 1	CCB87	-106	-0.0440	-7471B	PPB	09 Dec 2021 16:12:35	U	SMPL	
PB141284BS - 1	LCSS	21615	3.9080	-7471B	PPB	09 Dec 2021 16:17:32	U	SMPL	
M4966-05 - 1	P1-COMP	838	0.1278	-7471B	PPB	09 Dec 2021 16:19:48	U	SMPL	
M4966-14 - 1	P2-COMP	1021	0.1611	-7471B	PPB	09 Dec 2021 16:22:05	U	SMPL	
M4966-22 - 1	P3-COMP	34949	6.3340	-7471B	PPB	09 Dec 2021 16:24:21	U	SMPL	
M4966-29 - 1	P4-COMP	853	0.1305	-7471B	PPB	09 Dec 2021 16:26:37	U	SMPL	
M4967-01 - 1	TP-7S	1597	0.2658	-7471B	PPB	09 Dec 2021 16:28:55	U	SMPL	
M4967-01DUP - 1	TP-7SDUP	1833	0.3088	-7471B	PPB	09 Dec 2021 16:33:54	U	SMPL	
M4967-01MS - 1	TP-7SMS	19964	3.6076	-7471B	PPB	09 Dec 2021 16:36:10	U	SMPL	
M4967-01MSD - 1	TP-7SMSD	20445	3.6951	-7471B	PPB	09 Dec 2021 16:41:04	U	SMPL	
M4967-04 - 1	TP-7D	575	0.0799	-7471B	PPB	09 Dec 2021 16:43:19	U	SMPL	
CCV88 - 1	CCV88	28547	5.1692	-7471B	PPB	09 Dec 2021 16:45:35	U	SMPL	
CCB88 - 1	CCB88	-65	-0.0365	-7471B	PPB	09 Dec 2021 16:47:51	U	SMPL	
M4967-07 - 1	TP-8S	897	0.1385	-7471B	PPB	09 Dec 2021 16:50:08	U	SMPL	
M4967-10 - 1	TP-11S	1715	0.2873	-7471B	PPB	09 Dec 2021 16:52:24	U	SMPL	
M4967-13 - 1	TP-11D	1057	0.1676	-7471B	PPB	09 Dec 2021 16:54:39	U	SMPL	
M4973-01 - 1	HD-02-120821	797	0.1203	-7471B	PPB	09 Dec 2021 16:56:55	U	SMPL	
M4980-01 - 1	OK-03-120821	2302	0.3941	-7471B	PPB	09 Dec 2021 16:59:11	U	SMPL	
M4965-01MSD - 1	RED-PAINTCHIP-EXTERIORMSD	19704	3.5603	-7471B	PPB	09 Dec 2021 17:01:27	U	SMPL	
M4965-01LX5 - 1		80	-0.0102	-7471B	PPB	09 Dec 2021 17:03:43	U	SMPL	
M4965-01A - 1		18523	3.3454	-7471B	PPB	09 Dec 2021 17:06:01	U	SMPL	
M4967-01LX5 - 1		363	0.0413	-7471B	PPB	09 Dec 2021 17:08:17	U	SMPL	
M4967-01A - 1		19946	3.6043	-7471B	PPB	09 Dec 2021 17:10:34	U	SMPL	
CCV89 - 1	CCV89	25805	4.6703	-7471B	PPB	09 Dec 2021 17:12:50	U	SMPL	
CCB89 - 1	CCB89	-87	-0.0405	-7471B	PPB	09 Dec 2021 17:17:23	U	SMPL	