



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117730

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/09/2021	14:35
2	CAL2	1	Water	NA	NA	20.3	7.00	12/09/2021	14:36
3	CAL3	1	Water	NA	NA	20.4	10.03	12/09/2021	14:37
4	ICV	1	Water	NA	NA	20.5	7.00	12/09/2021	14:39
5	CCV1	1	Water	NA	NA	20.5	2.01	12/09/2021	14:40
6	M4993-01	1	Water	NA	NA	20.1	5.72	12/09/2021	15:40
7	M4993-02	1	Water	NA	NA	20.2	5.48	12/09/2021	15:41
8	M4993-03	1	Water	NA	NA	20.3	5.38	12/09/2021	15:45
9	M4998-01	1	Water	NA	NA	20.2	7.21	12/09/2021	15:46
10	M4998-01DUP	1	Water	NA	NA	20.3	7.22	12/09/2021	15:47
11	CCV2	1	Water	NA	NA	20.4	12.03	12/09/2021	15:50

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M4998 WorkList ID : 155021 Department : Wet-Chemistry Date : 12-09-2021 17:12:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/08/2021	Water	M4993-01	pH	Cool 4 deg C	STER07	D11	CELL-A	12/08/2021	9040C
12/08/2021	Water	M4993-02	pH	Cool 4 deg C	STER07	D11	CELLS-B/C	12/08/2021	9040C
12/08/2021	Water	M4993-03	pH	Cool 4 deg C	STER07	D11	CELL-D	12/08/2021	9040C
12/09/2021	Water	M4998-01	pH	Cool 4 deg C	PSEG03	H12	R24L1915LP	12/09/2021	9040C

Date/Time 12/09/21 15:30
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: JP (WC)

Date/Time 12/09/21 17:30
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: JP (WC)