

LB117733


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

39391



A= 0.0000e+000

B= 2.4912e-004

C= -7.1258e-002

Rho= 0.9988366

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.047	-0.047	98	0.000	98					
0.2	0.200	0.185	-0.015	1029	0.0 %	1029					-8
2.5	2.500	2.526	0.026	10424	0.0 %	10424					1
5.0	5.000	4.951	-0.049	20159	0.0 %	20159					-1
7.5	7.500	7.843	0.343	31770	0.0 %	31770					5
10.0	10.000	9.742	-0.258	39391	0.0 %	39391					-3

LB117733 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 10 Dec 2021 09:26:24

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	98	-	0.0000	7470A	PPB	10 Dec 2021 09:34:44	S	Std
0.2 - 1	50.2	1029	-	0.2000	7470A	PPB	10 Dec 2021 09:37:02	S	Std
2.5 - 1	52.5	10424	-	2.5000	7470A	PPB	10 Dec 2021 09:39:19	S	Std
5.0 - 1	55	20159	-	5.0000	7470A	PPB	10 Dec 2021 09:41:36	S	Std
7.5 - 1	57.5	31770	-	7.5000	7470A	PPB	10 Dec 2021 09:43:53	S	Std
10.0 - 1	58	39391	-	10.0000	7470A	PPB	10 Dec 2021 09:46:08	S	Std
ICV05 - 1	ICV05	16573	4.0575	-	7470A	PPB	10 Dec 2021 09:51:48	U	SMPL
ICB05 - 1	ICB05	75	-0.0526	-	7470A	PPB	10 Dec 2021 09:54:03	U	SMPL
CCV90 - 1	CCV90	19696	4.8355	-	7470A	PPB	10 Dec 2021 09:56:18	U	SMPL
CCB90 - 1	CCB90	65	-0.0551	-	7470A	PPB	10 Dec 2021 09:58:33	U	SMPL
CRA - 1	CRA	997	0.1771	-	7470A	PPB	10 Dec 2021 10:00:48	U	SMPL
HighStd - 1	HighStd	37663	9.3115	-	7470A	PPB	10 Dec 2021 10:03:03	U	SMPL
ChkStd - 1	ChkStd	28240	6.9640	-	7470A	PPB	10 Dec 2021 10:05:18	U	SMPL
PB141282BL - 1	PBW	185	-0.0252	-	7470A	PPB	10 Dec 2021 10:07:35	U	SMPL
PB141282BS - 1	LCSW	14657	3.5801	-	7470A	PPB	10 Dec 2021 10:09:51	U	SMPL
M4967-02 - 1	TP-7S	181	-0.0262	-	7470A	PPB	10 Dec 2021 10:12:07	U	SMPL
M4967-05 - 1	TP-7D	216	-0.0174	-	7470A	PPB	10 Dec 2021 10:14:23	U	SMPL
M4967-08 - 1	TP-8S	202	-0.0209	-	7470A	PPB	10 Dec 2021 10:16:40	U	SMPL
M4967-11 - 1	TP-11S	371	0.0212	-	7470A	PPB	10 Dec 2021 10:18:57	U	SMPL
M4967-14 - 1	TP-11D	344	0.0144	-	7470A	PPB	10 Dec 2021 10:21:12	U	SMPL
CCV91 - 1	CCV91	19620	4.8165	-	7470A	PPB	10 Dec 2021 10:23:28	U	SMPL
CCB91 - 1	CCB91	78	-0.0518	-	7470A	PPB	10 Dec 2021 10:25:43	U	SMPL
M4967-14DUP - 1	TP-11DDUP	367	0.0202	-	7470A	PPB	10 Dec 2021 10:27:59	U	SMPL
M4967-14MS - 1	TP-11DMS	15709	3.8422	-	7470A	PPB	10 Dec 2021 10:30:14	U	SMPL
M4967-14MSD - 1	TP-11DMSD	15599	3.8148	-	7470A	PPB	10 Dec 2021 10:32:30	U	SMPL
PB141282TB - 1	PB141262TB	212	-0.0184	-	7470A	PPB	10 Dec 2021 10:34:46	U	SMPL
M4967-14LX5 - 1		135	-0.0376	-	7470A	PPB	10 Dec 2021 10:37:02	U	SMPL
M4967-14A - 1		15754	3.8534	-	7470A	PPB	10 Dec 2021 10:39:18	U	SMPL
CCV92 - 1	CCV92	19637	4.8208	-	7470A	PPB	10 Dec 2021 10:41:34	U	SMPL
CCB92 - 1	CCB92	64	-0.0553	-	7470A	PPB	10 Dec 2021 10:43:51	U	SMPL