



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB117737

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	12/10/2021	13:55
2	M4998-01		1	10	100.00	>212.0	>212.0	12/10/2021	14:20
3	M4998-01DUP		1	10	100.00	>212.0	>212.0	12/10/2021	14:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB117737

WorkList Name : fp121021 WorkList ID : 155044 Department : Wet-Chemistry Date : 12-10-2021 09:52:58

Due Date	Matrix	Sample	Test	Flash Point	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/14/2021	Water	M4998-01	✓	Flash Point	Cool 4 deg C	PSEG03	H12	R24L1915LP	12/09/2021	1010B

Date/Time 12/10/21 13:30
Raw Sample Received by: KS (WC)
Raw Sample Relinquished by: RJ CEXA-1067

Date/Time 12/10/21 14:50
Raw Sample Received by: RJ CEXA-1067
Raw Sample Relinquished by: KS (WC)