


LB117739

7471B

INSTRUMENT ID: CV1

 μ Abs.:

50771



Linear

A= 0.0000e+000

B= 1.9586e-004

C= -5.2876e-002

Rho= 0.9996253

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/b
0.0	0.000	-0.045	-0.045	41	0.000	41					-
0.2	0.200	0.163	-0.037	1101	0.0 %	1101					-19
2.5	2.500	2.473	-0.027	12897	0.0 %	12897					-1
5.0	5.000	5.213	0.213	26884	0.0 %	26884					4
7.5	7.500	7.505	0.005	38590	0.0 %	38590					0
10.0	10.000	9.891	-0.109	50771	0.0 %	50771					-1

LB117739 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 10 Dec 2021 11:45:42

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	41	-	0.0000	7471B	PPB	10 Dec 2021 11:48:56	S	Std
0.2 - 1	50.2	1101	-	0.2000	7471B	PPB	10 Dec 2021 11:51:13	S	Std
2.5 - 1	52.1	12897	-	2.5000	7471B	PPB	10 Dec 2021 11:53:31	S	Std
5.0 - 1	55	26884	-	5.0000	7471B	PPB	10 Dec 2021 11:55:49	S	Std
7.5 - 1	57.5	38590	-	7.5000	7471B	PPB	10 Dec 2021 11:58:06	S	Std
10.0 - 1	511	50771	-	10.0000	7471B	PPB	10 Dec 2021 12:02:54	S	Std
ICV06 - 1	ICV06	20240	3.9113	-	7471B	PPB	10 Dec 2021 12:06:03	U	SMPL
ICB06 - 1	ICB06	-40	-0.0607	-	7471B	PPB	10 Dec 2021 12:08:17	U	SMPL
CCV93 - 1	CCV93	27520	5.3372	-	7471B	PPB	10 Dec 2021 12:10:33	U	SMPL
CCB93 - 1	CCB93	-52	-0.0631	-	7471B	PPB	10 Dec 2021 12:12:48	U	SMPL
CRA - 1	CRA	1070	0.1567	-	7471B	PPB	10 Dec 2021 12:15:03	U	SMPL
HighStd - 1	HighStd	50425	9.8233	-	7471B	PPB	10 Dec 2021 12:17:18	U	SMPL
ChkStd - 1	ChkStd	34677	6.7389	-	7471B	PPB	10 Dec 2021 12:19:34	U	SMPL
PB141321BL - 1	PBS	-25	-0.0578	-	7471B	PPB	10 Dec 2021 12:21:51	U	SMPL
PB141321BS - 1	LCSS	19226	3.7127	-	7471B	PPB	10 Dec 2021 12:24:08	U	SMPL
M4989-07 - 1	P5-COMP	3742	0.6800	-	7471B	PPB	10 Dec 2021 12:31:18	U	SMPL
M4989-07DUP - 1	P5-COMPDUP	3490	0.6307	-	7471B	PPB	10 Dec 2021 12:33:35	U	SMPL
M4989-07MS - 1	P5-COMPMS	22707	4.3945	-	7471B	PPB	10 Dec 2021 12:35:51	U	SMPL
M4989-07MSD - 1	P5-COMPMSD	22521	4.3581	-	7471B	PPB	10 Dec 2021 12:38:08	U	SMPL
M4989-15 - 1	P6-COMP	724	0.0889	-	7471B	PPB	10 Dec 2021 12:40:24	U	SMPL
CCV94 - 1	CCV94	26065	5.0522	-	7471B	PPB	10 Dec 2021 12:42:41	U	SMPL
CCB94 - 1	CCB94	-31	-0.0589	-	7471B	PPB	10 Dec 2021 12:44:56	U	SMPL
M4990-13 - 1	R2-COMP	11395	2.1789	-	7471B	PPB	10 Dec 2021 12:47:12	U	SMPL
M4990-15 - 1	R1-COMP	1529	0.2466	-	7471B	PPB	10 Dec 2021 12:49:27	U	SMPL
M4990-17 - 1	R3-COMP	8313	1.5753	-	7471B	PPB	10 Dec 2021 12:51:43	U	SMPL
M4989-07LX5 - 1		584	0.0615	-	7471B	PPB	10 Dec 2021 12:57:38	U	SMPL
M4989-07A - 1		23424	4.5349	-	7471B	PPB	10 Dec 2021 12:59:54	U	SMPL
CCV95 - 1	CCV95	25431	4.9280	-	7471B	PPB	10 Dec 2021 13:02:10	U	SMPL
CCB95 - 1	CCB95	-37	-0.0601	-	7471B	PPB	10 Dec 2021 13:04:27	U	SMPL