



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Iwona

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB117740

Reagent/Standard	Lot/Log #
Residual chlorine LOD, 0.05PPM	WP94323
Residual chlorine ICV-LCS, 0.4PPM	WP94322
Chlorine Calibration std, 0.1ppm	WP94315
Chlorine Calibration std, 0.2ppm	WP94317
Chlorine Calibration std, 0.8ppm	WP94320
Chlorine Calibration std, 0.0ppm	WP94314
Chlorine Calibration std, 1.6ppm	WP94321
Residual Chlorine LOQ 0.1ppm	
Residual Chlorine Calibration and CCV std, 0	WP94318

Intercept: 0.0043

Slope: 0.9498

Regression: 0.999902

Seq	Lab ID	True Val (mg/l)	DF	Initial Reading	Final Reading	Difference	Result (mg/l)	%D	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.010	0.010	0.01		12/10/2021	13:08
2	CAL2	0.1	1	0.000	0.100	0.100	0.10	-1	12/10/2021	13:08
3	CAL3	0.2	1	0.000	0.190	0.190	0.20	2	12/10/2021	13:09
4	CAL4	0.4	1	0.000	0.390	0.390	0.41	-1.5	12/10/2021	13:09
5	CAL5	0.8	1	0.000	0.750	0.750	0.79	1.9	12/10/2021	13:10
6	CAL6	1.6	1	0.000	1.530	1.530	1.61	-0.4	12/10/2021	13:10



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Iwona

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB117740

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.3800	0.3800	0.3960	12/10/2021	13:11
2	ICB				1	0.0000	0.0000	0.0000	-0.0050	12/10/2021	13:11
3	CCV1			0.4	1	0.0000	0.4000	0.4000	0.4170	12/10/2021	13:12
4	CCB1				1	0.0000	0.0000	0.0000	-0.0050	12/10/2021	13:12
5	LB117740BL	50	50		1	0.0000	0.0100	0.0100	0.0060	12/10/2021	13:13
6	LB117740BS	50	50	0.4	1	0.0000	0.3800	0.3800	0.3960	12/10/2021	13:13
7	M4068-09	50	50		1	0.0000	0.0500	0.0500	0.0480	12/10/2021	13:14
8	CCV2			0.4	1	0.0000	0.4000	0.4000	0.4170	12/10/2021	13:14
9	CCB2				1	0.0000	0.0100	0.0100	0.0060	12/10/2021	13:15