

LB117742


7470A

INSTRUMENT ID: CV4

Linear

 μ Abs.:

39846



A= 0.0000e+000

B= 2.5340e-004

C= -4.6892e-002

Rho= 0.9999343

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.000	-0.000	184	0.000	184					
0.2	0.200	0.212	0.012	1023	0.0 %	1023					
2.5	2.500	2.496	-0.004	10035	0.0 %	10035					
5.0	5.000	5.027	0.027	20024	0.0 %	20024					
7.5	7.500	7.415	-0.085	29445	0.0 %	29445					
10.0	10.000	10.050	0.050	39846	0.0 %	39846					

100
6
0
1
-1
)

LB117742 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 10 Dec 2021 14:58:27

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	184	-	0.0000	7470A	PPB	10 Dec 2021 15:02:21	S	Std
0.2 - 1	50.2	1023	-	0.2000	7470A	PPB	10 Dec 2021 15:04:38	S	Std
2.5 - 1	52.5	10035	-	2.5000	7470A	PPB	10 Dec 2021 15:06:55	S	Std
5.0 - 1	55	20024	-	5.0000	7470A	PPB	10 Dec 2021 15:09:12	S	Std
7.5 - 1	57.5	29445	-	7.5000	7470A	PPB	10 Dec 2021 15:11:29	S	Std
10.0 - 1	510	39846	-	10.0000	7470A	PPB	10 Dec 2021 15:16:15	S	Std
ICV07 - 1	ICV07	16031	4.0154	-	7470A	PPB	10 Dec 2021 15:18:53	U	SMPL
ICB07 - 1	ICB07	284	0.0251	-	7470A	PPB	10 Dec 2021 15:21:08	U	SMPL
CCV96 - 1	CCV96	19563	4.9104	-	7470A	PPB	10 Dec 2021 15:23:22	U	SMPL
CCB96 - 1	CCB96	289	0.0263	-	7470A	PPB	10 Dec 2021 15:25:37	U	SMPL
CRA - 1	CRA	973	0.1997	-	7470A	PPB	10 Dec 2021 15:27:52	U	SMPL
HighStd - 1	HighStd	38426	9.6903	-	7470A	PPB	10 Dec 2021 15:30:07	U	SMPL
ChkStd - 1	ChkStd	26796	6.7433	-	7470A	PPB	10 Dec 2021 15:32:22	U	SMPL
PB141318BL - 1	PBW	117	-0.0172	-	7470A	PPB	10 Dec 2021 15:34:39	U	SMPL
PB141318BS - 1	LCSW	16055	4.0215	-	7470A	PPB	10 Dec 2021 15:36:54	U	SMPL
M4990-14 - 1	R2-COMP	190	0.0013	-	7470A	PPB	10 Dec 2021 15:39:11	U	SMPL
M4990-16 - 1	R1-COMP	182	-0.0008	-	7470A	PPB	10 Dec 2021 15:41:28	U	SMPL
M4990-18 - 1	R3-COMP	189	0.0010	-	7470A	PPB	10 Dec 2021 15:43:45	U	SMPL
M4990-18DUP - 1	R3-COMPDUP	219	0.0086	-	7470A	PPB	10 Dec 2021 15:51:08	U	SMPL
M4990-18MS - 1	R3-COMPMS	16969	4.2531	-	7470A	PPB	10 Dec 2021 15:53:24	U	SMPL
CCV97 - 1	CCV97	19825	4.9768	-	7470A	PPB	10 Dec 2021 15:55:39	U	SMPL
CCB97 - 1	CCB97	287	0.0258	-	7470A	PPB	10 Dec 2021 15:57:54	U	SMPL
M4990-18MSD - 1	R3-COMPMSD	15449	3.8679	-	7470A	PPB	10 Dec 2021 16:00:09	U	SMPL
PB141319BL - 1	PBW	224	0.0099	-	7470A	PPB	10 Dec 2021 16:02:25	U	SMPL
PB141319BS - 1	LCSW	14218	3.5560	-	7470A	PPB	10 Dec 2021 16:04:41	U	SMPL
M4998-03 - 1	R24L1915LP	-52	-0.0601	-	7470A	PPB	10 Dec 2021 16:06:57	U	SMPL
M4998-03DUP - 1	R24L1915LPDUP	55	-0.0330	-	7470A	PPB	10 Dec 2021 16:09:12	U	SMPL
M4998-03MS - 1	R24L1915LPMS	13899	3.4751	-	7470A	PPB	10 Dec 2021 16:23:31	U	SMPL
M4998-03MSD - 1	R24L1915LPMSD	13739	3.4346	-	7470A	PPB	10 Dec 2021 16:25:48	U	SMPL
PB141320BL - 1	PBW	-63	-0.0629	-	7470A	PPB	10 Dec 2021 16:28:05	U	SMPL
PB141320BS - 1	LCSW	15329	3.8375	-	7470A	PPB	10 Dec 2021 16:33:08	U	SMPL
M4998-01 - 1	R24L1915LP	9	-0.0446	-	7470A	PPB	10 Dec 2021 16:35:23	U	SMPL
CCV98 - 1	CCV98	18720	4.6968	-	7470A	PPB	10 Dec 2021 16:37:39	U	SMPL
CCB98 - 1	CCB98	227	0.0106	-	7470A	PPB	10 Dec 2021 16:39:55	U	SMPL
M4998-01DUP - 1	R24L1915LPDUP	67	-0.0299	-	7470A	PPB	10 Dec 2021 16:42:11	U	SMPL
M4998-01MS - 1	R24L1915LPMS	15068	3.7714	-	7470A	PPB	10 Dec 2021 16:47:00	U	SMPL
M4998-01MSD - 1	R24L1915LPMSD	15162	3.7952	-	7470A	PPB	10 Dec 2021 16:49:15	U	SMPL
PB141318TB - 1	PB141286TB	117	-0.0172	-	7470A	PPB	10 Dec 2021 16:51:31	U	SMPL
PB141319TB - 1	PB141297TB	163	-0.0056	-	7470A	PPB	10 Dec 2021 16:53:47	U	SMPL
CCV99 - 1	CCV99	18538	4.6507	-	7470A	PPB	10 Dec 2021 16:56:03	U	SMPL
CCB99 - 1	CCB99	140	-0.0114	-	7470A	PPB	10 Dec 2021 16:58:20	U	SMPL
M4990-18LX5 - 1		158	-0.0069	-	7470A	PPB	10 Dec 2021 17:00:37	U	SMPL
M4990-18A - 1		14193	3.5496	-	7470A	PPB	10 Dec 2021 17:02:54	U	SMPL
M4998-03LX5 - 1		71	-0.0289	-	7470A	PPB	10 Dec 2021 17:05:09	U	SMPL
M4998-03A - 1		15236	3.8139	-	7470A	PPB	10 Dec 2021 17:07:24	U	SMPL
M4998-01LX5 - 1		117	-0.0172	-	7470A	PPB	10 Dec 2021 17:09:40	U	SMPL
M4998-01A - 1		15646	3.9178	-	7470A	PPB	10 Dec 2021 17:11:56	U	SMPL
CCV01 - 1	CCV01	20383	5.1182	-	7470A	PPB	10 Dec 2021 17:14:11	U	SMPL
CCB01 - 1	CCB01	42	-0.0362	-	7470A	PPB	10 Dec 2021 17:16:27	U	SMPL