



PERCENT SOLID

Supervisor: ketankumar

Analyst: jignesh

Date: 12/11/2021

OVENTEMP IN Celsius(°C): 107

Time IN: 15:30

In Date: 12/10/2021

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:22

Out Date: 12/11/2021

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB117743

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M5023-01	BGKX7	1	1.17	8.48	9.65	7.21	71.2	
M5023-02	BGKX8	2	1.15	8.80	9.95	9.14	90.8	
M5023-03	BGKX9	3	1.18	8.53	9.71	8.45	85.2	
M5023-04	BGKY0	4	1.17	8.81	9.98	8.84	87.1	
M5023-05	BGKY1	5	1.17	8.65	9.82	8.58	85.7	
M5023-06	BGKY2	6	1.15	8.62	9.77	7.4	72.5	
M5023-07	BGKY3	7	1.17	8.60	9.77	7.36	72.0	
M5023-08	BGKY4	8	1.16	8.77	9.93	7.76	75.3	
M5023-10	BGKY6	9	1.13	8.39	9.52	8.7	90.2	
M5023-11	BGKY7	10	1.13	8.84	9.97	9.2	91.3	
M5023-12	BGKY8	11	1.12	8.51	9.63	8.68	88.8	
M5023-13	BGKY9	12	1.11	8.62	9.73	8.11	81.2	
M5023-14	BGKZ0	13	1.18	8.78	9.96	8.67	85.3	
M5023-15	BGKZ2	14	1.11	8.41	9.52	7.1	71.2	
M5023-16	BGKZ4	15	1.15	8.82	9.97	7.41	71.0	
M5023-17	BGKZ5	16	1.11	8.83	9.94	9.44	94.3	
M5023-18	BGKZ6	17	1.13	8.80	9.93	9.06	90.1	
M5023-19	BGKZ7	18	1.17	8.81	9.98	8.4	82.1	
M5023-20	BGKZ8	19	1.19	8.51	9.7	7.46	73.7	
M5023-21	VHBLK001	20	1.00	1.00	2.00	2.00	100.0	VHBLK

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$