

LB117764


7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

39073



A= 0.0000e+000

B= 2.5924e-004

C= -3.1262e-002

Rho= 0.9997390

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/01)
0.0	0.000	0.026	0.026	219	0.000	219					-16
0.2	0.200	0.168	-0.032	768	0.0 %	768					2
2.5	2.500	2.558	0.058	9987	0.0 %	9987					0
5.0	5.000	5.017	0.017	19473	0.0 %	19473					-2
7.5	7.500	7.334	-0.166	28409	0.0 %	28409					1
10.0	10.000	10.098	0.098	39073	0.0 %	39073					

LB117764 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 13 Dec 2021 12:13:32						
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type	
0.0 - 1	50	219	-	0.0000	7470A	PPB	13 Dec 2021 12:34:15	S	Std	
0.2 - 1	51.2	768	-	0.2000	7470A	PPB	13 Dec 2021 12:36:32	S	Std	
2.5 - 1	52.5	9987	-	2.5000	7470A	PPB	13 Dec 2021 12:38:49	S	Std	
5.0 - 1	55	19473	-	5.0000	7470A	PPB	13 Dec 2021 12:41:06	S	Std	
7.5 - 1	57.5	28409	-	7.5000	7470A	PPB	13 Dec 2021 12:46:01	S	Std	
10.0 - 1	510	39073	-	10.0000	7470A	PPB	13 Dec 2021 12:50:54	S	Std	
ICV08 - 1	ICV08	15834	4.0736	-	7470A	PPB	13 Dec 2021 12:53:33	U	SMPL	
ICB08 - 1	ICB08	89	-0.0082	-	7470A	PPB	13 Dec 2021 12:55:48	U	SMPL	
CCV02 - 1	CCV02	20522	5.2890	-	7470A	PPB	13 Dec 2021 12:58:03	U	SMPL	
CCB02 - 1	CCB02	80	-0.0105	-	7470A	PPB	13 Dec 2021 13:00:18	U	SMPL	
CRA - 1	CRA	746	0.1621	-	7470A	PPB	13 Dec 2021 13:02:33	U	SMPL	
HighStd - 1	HighStd	38822	10.0331	-	7470A	PPB	13 Dec 2021 13:04:48	U	SMPL	
ChkStd - 1	ChkStd	27021	6.9738	-	7470A	PPB	13 Dec 2021 13:07:03	U	SMPL	
PB141361BL - 1	PBW	-65	-0.0481	-	7470A	PPB	13 Dec 2021 13:09:20	U	SMPL	
PB141361BS - 1	LCSW	13779	3.5409	-	7470A	PPB	13 Dec 2021 13:11:35	U	SMPL	
M4966-06 - 1	P1-COMP	-12	-0.0344	-	7470A	PPB	13 Dec 2021 13:13:51	U	SMPL	
M4966-06DUP - 1	P1-COMPDUP	34	-0.0224	-	7470A	PPB	13 Dec 2021 13:16:08	U	SMPL	
M4966-06MS - 1	P1-COMPMS	15180	3.9041	-	7470A	PPB	13 Dec 2021 13:18:25	U	SMPL	
M4966-06MSD - 1	P1-COMPMSD	14748	3.7921	-	7470A	PPB	13 Dec 2021 13:20:41	U	SMPL	
M4966-15 - 1	P2-COMP	-14	-0.0349	-	7470A	PPB	13 Dec 2021 13:22:57	U	SMPL	
CCV03 - 1	CCV03	20489	5.2804	-	7470A	PPB	13 Dec 2021 13:25:12	U	SMPL	
CCB03 - 1	CCB03	81	-0.0103	-	7470A	PPB	13 Dec 2021 13:27:27	U	SMPL	
M4966-23 - 1	P3-COMP	25	-0.0248	-	7470A	PPB	13 Dec 2021 13:29:43	U	SMPL	
M4966-30 - 1	P4-COMP	93	-0.0072	-	7470A	PPB	13 Dec 2021 13:31:58	U	SMPL	
M4989-08 - 1	P5-COMP	44	-0.0199	-	7470A	PPB	13 Dec 2021 13:34:13	U	SMPL	
M4989-16 - 1	P6-COMP	149	0.0074	-	7470A	PPB	13 Dec 2021 13:36:29	U	SMPL	
PB141362BL - 1	PBW	112	-0.0022	-	7470A	PPB	13 Dec 2021 13:38:44	U	SMPL	
PB141362BS - 1	LCSW	13083	3.3604	-	7470A	PPB	13 Dec 2021 13:41:01	U	SMPL	
M5017-10 - 1	TP1	81	-0.0103	-	7470A	PPB	13 Dec 2021 13:43:17	U	SMPL	
M5017-20 - 1	TP2	92	-0.0074	-	7470A	PPB	13 Dec 2021 13:45:34	U	SMPL	
M5017-30 - 1	TP3	109	-0.0030	-	7470A	PPB	13 Dec 2021 13:47:51	U	SMPL	
M5017-40 - 1	SP1A	297	0.0457	-	7470A	PPB	13 Dec 2021 13:50:06	U	SMPL	
CCV04 - 1	CCV04	20419	5.2623	-	7470A	PPB	13 Dec 2021 13:52:22	U	SMPL	
CCB04 - 1	CCB04	146	0.0066	-	7470A	PPB	13 Dec 2021 13:54:37	U	SMPL	
M5017-50 - 1	SP1-B	96	-0.0064	-	7470A	PPB	13 Dec 2021 13:56:53	U	SMPL	
M5017-60 - 1	SP2-A	104	-0.0043	-	7470A	PPB	13 Dec 2021 13:59:08	U	SMPL	
M5018-10 - 1	SP2B	82	-0.0100	-	7470A	PPB	13 Dec 2021 14:01:24	U	SMPL	
M5018-20 - 1	TP4	92	-0.0074	-	7470A	PPB	13 Dec 2021 14:03:39	U	SMPL	
M5018-30 - 1	TP5	67	-0.0139	-	7470A	PPB	13 Dec 2021 14:05:55	U	SMPL	
M5018-40 - 1	TP6	123	0.0006	-	7470A	PPB	13 Dec 2021 14:08:12	U	SMPL	
M5018-50 - 1	TP7	119	-0.0004	-	7470A	PPB	13 Dec 2021 14:10:28	U	SMPL	
M5018-60 - 1	TP8	122	0.0004	-	7470A	PPB	13 Dec 2021 14:12:45	U	SMPL	
M5018-70 - 1	TP9	113	-0.0020	-	7470A	PPB	13 Dec 2021 14:15:02	U	SMPL	
M5022-10 - 1	TP10	130	0.0024	-	7470A	PPB	13 Dec 2021 14:17:18	U	SMPL	
CCV05 - 1	CCV05	20474	5.2765	-	7470A	PPB	13 Dec 2021 14:19:33	U	SMPL	
CCB05 - 1	CCB05	192	0.0185	-	7470A	PPB	13 Dec 2021 14:21:49	U	SMPL	
M5022-20 - 1	TP11	99	-0.0056	-	7470A	PPB	13 Dec 2021 14:24:04	U	SMPL	
M5022-30 - 1	TP12	99	-0.0056	-	7470A	PPB	13 Dec 2021 14:26:20	U	SMPL	
M5022-40 - 1	TP13	104	-0.0043	-	7470A	PPB	13 Dec 2021 14:28:35	U	SMPL	
M5022-50 - 1	TP14	95	-0.0066	-	7470A	PPB	13 Dec 2021 14:30:51	U	SMPL	
M5022-60 - 1	TP15	92	-0.0074	-	7470A	PPB	13 Dec 2021 14:33:06	U	SMPL	
M5022-60DUP - 1	TP15DUP	83	-0.0097	-	7470A	PPB	13 Dec 2021 14:35:23	U	SMPL	
M5022-60MS - 1	TP15MS	15126	3.8901	-	7470A	PPB	13 Dec 2021 14:37:39	U	SMPL	
M5022-60MSD - 1	TP15MSD	15541	3.9977	-	7470A	PPB	13 Dec 2021 14:39:56	U	SMPL	
CCV06 - 1	CCV06	19144	4.9317	-	7470A	PPB	13 Dec 2021 14:42:13	U	SMPL	
CCB06 - 1	CCB06	183	0.0162	-	7470A	PPB	13 Dec 2021 14:44:29	U	SMPL	
M4966-06LX5 - 1		7	-0.0294	-	7470A	PPB	13 Dec 2021 14:46:45	U	SMPL	
M4966-06A - 1		16099	4.1423	-	7470A	PPB	13 Dec 2021 14:49:01	U	SMPL	
M5022-60LX5 - 1		-17	-0.0357	-	7470A	PPB	13 Dec 2021 14:51:16	U	SMPL	
M5022-60A - 1		16116	4.1467	-	7470A	PPB	13 Dec 2021 14:53:31	U	SMPL	
PB141361TB - 1		8	-0.0292	-	7470A	PPB	13 Dec 2021 14:55:47	U	SMPL	
PB141362TB - 1		122	0.0004	-	7470A	PPB	13 Dec 2021 14:58:03	U	SMPL	
CCV07 - 1	CCV07	19985	5.1497	-	7470A	PPB	13 Dec 2021 15:00:19	U	SMPL	
CCB07 - 1	CCB07	117	-0.0009	-	7470A	PPB	13 Dec 2021 15:02:35	U	SMPL	