



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117772

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/13/2021	16:10
2	CAL2	1	Water	NA	NA	20.2	7.00	12/13/2021	16:11
3	CAL3	1	Water	NA	NA	20.2	10.03	12/13/2021	16:12
4	ICV	1	Water	NA	NA	20.3	7.02	12/13/2021	16:15
5	CCV1	1	Water	NA	NA	20.4	2.01	12/13/2021	16:16
6	M5072-01	1	Water	NA	NA	20.3	7.30	12/13/2021	16:20
7	M5072-01DUP	1	Water	NA	NA	20.4	7.31	12/13/2021	16:21
8	M5072-02	1	Water	NA	NA	20.1	6.84	12/13/2021	16:24
9	CCV2	1	Water	NA	NA	20.5	12.03	12/13/2021	16:25

WB 11772

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M5072 W WorkList ID : 155150 Department : Wet-Chemistry Date : 12-13-2021 16:00:33

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/13/2021	Water	M5072-01	pH	Cool 4 deg C	PSEG03	H22	T2263	12/13/2021	9040C
12/13/2021	Water	M5072-02	pH	Cool 4 deg C	PSEG03	H22	T2264	12/13/2021	9040C

Date/Time 12-13-21 16:05
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: JP (SM)

Date/Time 12-13-21 16:40
Raw Sample Received by: JP (SM)
Raw Sample Relinquished by: JP (WC)