



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB117775

Reagent/Standard	Lot/Loq #
TURBIDITY LOD STD, 0.5NTU	WP94372
Turbidity Calibration std, 0NTU	WP94370
Turbidity Calibration std, 20NTU	WP94365
Turbidity Calibration std, 80NTU	WP94363
Turbidity Calibration std, 40NTU	WP94364
Turbidity Calibration std, 200NTU	WP94371
Turbidity Calibration std, 1NTU	WP94368
Turbidity Calibration std, 5NTU	WP94367
Turbidity LOQ std, 1.0NTU	WP94373
Turbidity Calibration - CCV std, 10 NTU	WP94366

Intercept: 0.2435

Slope: 0.8525

Regression: 0.9996

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.162	-0.10		12/13/2021	14:42
2	CAL2	1	1	1.050	0.95	5.4	12/13/2021	14:42
3	CAL3	5	1	4.300	4.76	4.8	12/13/2021	14:43
4	CAL4	10	1	9.440	10.79	-7.9	12/13/2021	14:43
5	CAL5	20	1	16.900	19.54	2.3	12/13/2021	14:44
6	CAL6	40	1	34.400	40.07	-0.2	12/13/2021	14:44



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB117775

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	8.380	9.544	12/13/2021	14:45
2	ICB		1	0.154	-0.105	12/13/2021	14:45
3	CCV1	10	1	8.640	9.849	12/13/2021	14:46
4	CCB1		1	0.187	-0.066	12/13/2021	14:46
5	LB117775BL		1	0.188	-0.065	12/13/2021	14:47
6	M4068-10		1	0.665	0.494	12/13/2021	14:47
7	M4068-11		1	1.260	1.192	12/13/2021	14:48
8	CCV2	10	1	8.460	9.638	12/13/2021	14:48
9	CCB2		1	0.175	-0.080	12/13/2021	14:49