



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB117776

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	12/13/2021	15:15
2	CAL2	1	Water	NA	NA	20.4	7.00	12/13/2021	15:16
3	CAL3	1	Water	NA	NA	20.4	10.03	12/13/2021	15:17
4	ICV	1	Water	NA	NA	20.5	7.01	12/13/2021	15:20
5	CCV1	1	Water	NA	NA	20.5	2.01	12/13/2021	15:22
6	M5065-02	1	Solid	20.03	20	20.2	9.20	12/13/2021	15:25
7	M5067-02	1	Solid	20.04	20	20.6	10.41	12/13/2021	15:28
8	M5067-05	1	Solid	20.05	20	20.6	9.62	12/13/2021	15:35
9	M5067-05DUP	1	Solid	20.04	20	20.7	9.63	12/13/2021	15:36
10	CCV2	1	Water	NA	NA	20.5	12.02	12/13/2021	15:40

WORKLIST(Hardcopy Internal Chain)

VB 117776

WorkList Name : CORROSIONITY M5066 WorkList ID : 155121 Department : Wet-Chemistry Date : 12-13-2021 13:49:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/10/2021	Solid	M5065-02	Corrosivity	Cool 4 deg C	RMJE02	H22	UV-1WC	12/10/2021	9045D
12/13/2021	Solid	M5067-02	Corrosivity	Cool 4 deg C	PSEG03	H22	BASE	12/13/2021	9045D
12/13/2021	Solid	M5067-05	Corrosivity	Cool 4 deg C	PSEG03	H22	BRICK	12/13/2021	9045D

Date/Time 12/13/21 15:10
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: JP (WC)

Date/Time 12/13/21 16:00
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: JP (WC)