



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB117791

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	12/14/2021	12:10
2	M5072-03		1	12	100.00	>212.0	>212.0	12/14/2021	12:35
3	M5072-03DUP		1	12	100.00	>212.0	>212.0	12/14/2021	12:58

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB117791

WorkList Name : fp121321 WorkList ID : 155103 Department : Wet-Chemistry Date : 12-13-2021 09:10:23

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/20/2021	Water	M5072-03	Flash Point	Cool 4 deg C	PSEG03	H22	2263-64	12/13/2021	1010B

Date/Time 12/14/21 11:45
 Raw Sample Received by: KS(wc)
 Raw Sample Relinquished by: JP(wc)

Date/Time 12/14/21 13:05
 Raw Sample Received by: JP(wc)
 Raw Sample Relinquished by: KS(wc)