



Analytical Summary Report

Analysis Method: 9045D

Analyst By : ketankumar

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117793

Slope : 98.8

BalanceID:

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/14/2021	15:50
2	CAL2	1	Water	NA	NA	20.5	7.01	12/14/2021	15:52
3	CAL3	1	Water	NA	NA	20.3	10.02	12/14/2021	15:53
4	ICV	1	Water	NA	NA	20.4	7.02	12/14/2021	15:55
5	CCV1	1	Water	NA	NA	20.4	2.00	12/14/2021	15:57
6	M5088-01	1	Solid	20	20	21.8	7.07	12/14/2021	16:00
7	M5088-01DUP	1	Solid	20	20	21.8	7.08	12/14/2021	16:02
8	M5088-02	1	Solid	20	20	21.4	7.53	12/14/2021	16:05
9	M5088-03	1	Solid	20	20	20.6	6.63	12/14/2021	16:08
10	M5088-04	1	Solid	20	20	20.3	6.75	12/14/2021	16:10
11	M5088-05	1	Solid	20	20	20.6	6.42	12/14/2021	16:12
12	M5088-06	1	Solid	20	20	20.3	6.20	12/14/2021	16:15
13	M5088-07	1	Solid	20	20	20.1	6.31	12/14/2021	16:18
14	M5088-08	1	Solid	20	20	20.6	6.38	12/14/2021	16:20
15	CCV2	1	Water	NA	NA	20.4	12.01	12/14/2021	16:22

117793

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M5088 WorkList ID : 155177 Department : Wet-Chemistry Date : 12-14-2021 15:40:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/14/2021	Solid	M5088-01	pH	Cool 4 deg C	NEWJ04	H31	IB-1-0.5	12/14/2021	9045D
12/14/2021	Solid	M5088-02	pH	Cool 4 deg C	NEWJ04	H31	IB-1-1.5	12/14/2021	9045D
12/14/2021	Solid	M5088-03	pH	Cool 4 deg C	NEWJ04	H31	IB-2-0.5	12/14/2021	9045D
12/14/2021	Solid	M5088-04	pH	Cool 4 deg C	NEWJ04	H31	IB-2-1.5	12/14/2021	9045D
12/14/2021	Solid	M5088-05	pH	Cool 4 deg C	NEWJ04	H31	IB-3-0.5	12/14/2021	9045D
12/14/2021	Solid	M5088-06	pH	Cool 4 deg C	NEWJ04	H31	IB-3-1.5	12/14/2021	9045D
12/14/2021	Solid	M5088-07	pH	Cool 4 deg C	NEWJ04	H31	IB-4-0.5	12/14/2021	9045D
12/14/2021	Solid	M5088-08	pH	Cool 4 deg C	NEWJ04	H31	IB-4-1.5	12/14/2021	9045D

Date/Time 12-14-21 15:45
 Raw Sample Received by: 701001
 Raw Sample Relinquished by: C. S. M.

Date/Time 12-14-21 17:00
 Raw Sample Received by: C. S. M.
 Raw Sample Relinquished by: 701001

16:25