

LB117794


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

57706



A= 0.0000e+000

B= 1.7459e-004

C= 1.2113e-002

Rho= 0.9998816

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.037	0.037	141	0.000	141					3
0.2	0.200	0.207	0.007	1114	0.0 %	1114					0
2.5	2.500	2.498	-0.002	14238	0.0 %	14238					1
5.0	5.000	4.969	-0.031	28394	0.0 %	28394					1
7.5	7.500	7.403	-0.097	42332	0.0 %	42332					1
10.0	10.000	10.087	0.087	57706	0.0 %	57706					1

LB117794 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 14 Dec 2021 16:28:25

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	141	-	0.0000	7471B	PPB	14 Dec 2021 16:30:19	S	Std
0.2 - 1	52	1114	-	0.2000	7471B	PPB	14 Dec 2021 16:32:36	S	Std
2.5 - 1	52.5	14238	-	2.5000	7471B	PPB	14 Dec 2021 16:34:55	S	Std
5.0 - 1	55	28394	-	5.0000	7471B	PPB	14 Dec 2021 16:37:11	S	Std
7.5 - 1	57.5	42332	-	7.5000	7471B	PPB	14 Dec 2021 16:39:29	S	Std
10.0 - 1	510	57706	-	10.0000	7471B	PPB	14 Dec 2021 16:41:47	S	Std
ICV11 - 1	ICV11	21523	3.7697	-	7471B	PPB	14 Dec 2021 16:44:32	U	SMPL
ICB11 - 1	ICB11	-78	-0.0015	-	7471B	PPB	14 Dec 2021 16:46:46	U	SMPL
CCV15 - 1	CCV15	27738	4.8548	-	7471B	PPB	14 Dec 2021 16:49:03	U	SMPL
CCB15 - 1	CCB15	82	0.0264	-	7471B	PPB	14 Dec 2021 16:51:20	U	SMPL
CRA - 1	CRA	1075	0.1998	-	7471B	PPB	14 Dec 2021 16:53:39	U	SMPL
HighStd - 1	HighStd	55681	9.7332	-	7471B	PPB	14 Dec 2021 16:55:53	U	SMPL
ChkStd - 1	ChkStd	38906	6.8046	-	7471B	PPB	14 Dec 2021 16:58:10	U	SMPL
PB141386BL - 1	PBS	33	0.0179	-	7471B	PPB	14 Dec 2021 17:03:01	U	SMPL
PB141386BS - 1	LCSS	26041	4.5585	-	7471B	PPB	14 Dec 2021 17:05:17	U	SMPL
M5017-09 - 1	TP1	2853	0.5102	-	7471B	PPB	14 Dec 2021 17:07:36	U	SMPL
M5017-19 - 1	TP2	4552	0.8068	-	7471B	PPB	14 Dec 2021 17:09:53	U	SMPL
M5017-29 - 1	TP3	2051	0.3702	-	7471B	PPB	14 Dec 2021 17:12:09	U	SMPL
M5017-39 - 1	SP1A	2352	0.4227	-	7471B	PPB	14 Dec 2021 17:14:28	U	SMPL
M5017-49 - 1	SP1-B	1832	0.3320	-	7471B	PPB	14 Dec 2021 17:16:44	U	SMPL
CCV16 - 1	CCV16	27840	4.8726	-	7471B	PPB	14 Dec 2021 17:19:03	U	SMPL
CCB16 - 1	CCB16	131	0.0350	-	7471B	PPB	14 Dec 2021 17:21:18	U	SMPL
M5017-59 - 1	SP2-A	4205	0.7462	-	7471B	PPB	14 Dec 2021 17:23:33	U	SMPL
M5018-09 - 1	SP2B	3645	0.6485	-	7471B	PPB	14 Dec 2021 17:25:48	U	SMPL
M5018-19 - 1	TP4	3162	0.5642	-	7471B	PPB	14 Dec 2021 17:28:06	U	SMPL
M5018-29 - 1	TP5	3413	0.6080	-	7471B	PPB	14 Dec 2021 17:30:22	U	SMPL
M5018-39 - 1	TP6	2899	0.5182	-	7471B	PPB	14 Dec 2021 17:32:37	U	SMPL
M5018-49 - 1	TP7	4182	0.7422	-	7471B	PPB	14 Dec 2021 17:34:55	U	SMPL
M5018-59 - 1	TP8	5329	0.9425	-	7471B	PPB	14 Dec 2021 17:37:11	U	SMPL
M5018-69 - 1	TP9	3277	0.5842	-	7471B	PPB	14 Dec 2021 17:39:28	U	SMPL
M5022-09 - 1	TP10	2348	0.4220	-	7471B	PPB	14 Dec 2021 17:41:45	U	SMPL
M5022-19 - 1	TP11	2334	0.4196	-	7471B	PPB	14 Dec 2021 17:44:03	U	SMPL
CCV17 - 1	CCV17	28367	4.9646	-	7471B	PPB	14 Dec 2021 17:46:19	U	SMPL
CCB17 - 1	CCB17	-15	0.0095	-	7471B	PPB	14 Dec 2021 17:48:35	U	SMPL
M5022-29 - 1	TP12	6263	1.1055	-	7471B	PPB	14 Dec 2021 17:50:50	U	SMPL
M5022-39 - 1	TP13	1661	0.3021	-	7471B	PPB	14 Dec 2021 17:53:08	U	SMPL
M5022-49 - 1	TP14	4401	0.7805	-	7471B	PPB	14 Dec 2021 17:55:23	U	SMPL
M5022-59 - 1	TP15	4935	0.8737	-	7471B	PPB	14 Dec 2021 17:57:40	U	SMPL
M5022-59DUP - 1	TP15DUP	5258	0.9301	-	7471B	PPB	14 Dec 2021 18:03:07	U	SMPL
M5022-59MS - 1	TP15MS	27379	4.7921	-	7471B	PPB	14 Dec 2021 18:05:25	U	SMPL
M5022-59MSD - 1	TP15MSD	27330	4.7836	-	7471B	PPB	14 Dec 2021 18:07:45	U	SMPL
M5022-59LX5 - 1		896	0.1685	-	7471B	PPB	14 Dec 2021 18:10:02	U	SMPL
M5022-59A - 1		27381	4.7925	-	7471B	PPB	14 Dec 2021 18:12:24	U	SMPL
CCV18 - 1	CCV18	28206	4.9365	-	7471B	PPB	14 Dec 2021 18:14:39	U	SMPL
CCB18 - 1	CCB18	2	0.0125	-	7471B	PPB	14 Dec 2021 18:16:55	U	SMPL