

LB117797


7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

41630



A= 0.0000e+000

B= 2.4167e-004

C= -7.1282e-002

Rho= 0.9997466

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.001	0.001	298	0.000	298					
0.2	0.200	0.216	0.016	1187	0.0 %	1187					
2.5	2.500	2.562	0.062	10897	0.0 %	10897					
5.0	5.000	4.834	-0.166	20297	0.0 %	20297					
7.5	7.500	7.598	0.098	31734	0.0 %	31734					
10.0	10.000	9.990	-0.010	41630	0.0 %	41630					

1
2
3
7
0

LB117797 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 Dec 2021 08:35:00

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	298	-	0.0000	7470A	PPB	15 Dec 2021 08:52:21	S	Std
0.2 - 1	50-2	1187	-	0.2000	7470A	PPB	15 Dec 2021 08:54:38	S	Std
2.5 - 1	52-5	10897	-	2.5000	7470A	PPB	15 Dec 2021 08:56:55	S	Std
5.0 - 1	55	20297	-	5.0000	7470A	PPB	15 Dec 2021 08:59:12	S	Std
7.5 - 1	57-5	31734	-	7.5000	7470A	PPB	15 Dec 2021 09:01:29	S	Std
10.0 - 1	510	41630	-	10.0000	7470A	PPB	15 Dec 2021 09:03:46	S	Std
ICV12 - 1	ICV12	16809	3.9910	-	7470A	PPB	15 Dec 2021 09:06:28	U	SMPL
ICB12 - 1	ICB12	447	0.0367	-	7470A	PPB	15 Dec 2021 09:08:43	U	SMPL
CCV19 - 1	CCV19	20758	4.9454	-	7470A	PPB	15 Dec 2021 09:10:58	U	SMPL
CCB19 - 1	CCB19	101	-0.0469	-	7470A	PPB	15 Dec 2021 09:13:13	U	SMPL
CRA - 1	CRA	1186	0.2153	-	7470A	PPB	15 Dec 2021 09:15:28	U	SMPL
HighStd - 1	HighStd	41862	10.0456	-	7470A	PPB	15 Dec 2021 09:17:43	U	SMPL
ChkStd - 1	ChkStd	29706	7.1078	-	7470A	PPB	15 Dec 2021 09:19:59	U	SMPL
PB141430BL - 1	PBW	380	0.0206	-	7470A	PPB	15 Dec 2021 09:22:15	U	SMPL
PB141430BS - 1	LCSW	15500	3.6746	-	7470A	PPB	15 Dec 2021 09:24:31	U	SMPL
M5029-02 - 1	PSC-124035	469	0.0421	-	7470A	PPB	15 Dec 2021 09:26:47	U	SMPL
M5029-04 - 1	PSC-124019	798	0.1216	-	7470A	PPB	15 Dec 2021 09:29:03	U	SMPL
M5065-02 - 1	UV-1WC	506	0.0510	-	7470A	PPB	15 Dec 2021 09:31:20	U	SMPL
M5067-02 - 1	BASE	509	0.0517	-	7470A	PPB	15 Dec 2021 09:33:37	U	SMPL
M5067-05 - 1	BRICK	501	0.0498	-	7470A	PPB	15 Dec 2021 09:35:53	U	SMPL
CCV20 - 1	CCV20	20895	4.9785	-	7470A	PPB	15 Dec 2021 09:38:08	U	SMPL
CCB20 - 1	CCB20	61	-0.0565	-	7470A	PPB	15 Dec 2021 09:40:23	U	SMPL
M5074-01 - 1	CS-1	503	0.0503	-	7470A	PPB	15 Dec 2021 09:42:39	U	SMPL
M5074-03 - 1	CS-2	502	0.0500	-	7470A	PPB	15 Dec 2021 09:44:54	U	SMPL
M5074-05 - 1	CS-3	479	0.0445	-	7470A	PPB	15 Dec 2021 09:47:10	U	SMPL
M5074-07 - 1	CS-4	501	0.0498	-	7470A	PPB	15 Dec 2021 09:49:25	U	SMPL
M5074-07DUP - 1	CS-4DUP	513	0.0527	-	7470A	PPB	15 Dec 2021 09:51:41	U	SMPL
M5074-07MS - 1	CS-4MS	15185	3.5985	-	7470A	PPB	15 Dec 2021 09:53:58	U	SMPL
M5074-07MSD - 1	CS-4MSD	15536	3.6833	-	7470A	PPB	15 Dec 2021 09:56:14	U	SMPL
PB141431BL - 1	PBW	475	0.0435	-	7470A	PPB	15 Dec 2021 09:58:32	U	SMPL
PB141431BS - 1	LCSW	15865	3.7629	-	7470A	PPB	15 Dec 2021 10:00:48	U	SMPL
CCV21 - 1	CCV21	20784	4.9516	-	7470A	PPB	15 Dec 2021 10:03:04	U	SMPL
CCB21 - 1	CCB21	72	-0.0539	-	7470A	PPB	15 Dec 2021 10:05:20	U	SMPL
M5072-03 - 1	2263-64	481	0.0450	-	7470A	PPB	15 Dec 2021 10:07:35	U	SMPL
M5072-03DUP - 1	2263-64DUP	461	0.0401	-	7470A	PPB	15 Dec 2021 10:09:50	U	SMPL
M5072-03MS - 1	2263-64MS	12004	2.8298	-	7470A	PPB	15 Dec 2021 10:12:06	U	SMPL
M5072-03MSD - 1	2263-64MSD	11854	2.7935	-	7470A	PPB	15 Dec 2021 10:14:21	U	SMPL
PB141430TB - 1	PB141363TB	494	0.0481	-	7470A	PPB	15 Dec 2021 10:16:38	U	SMPL
M5074-07LX5 - 1		100	-0.0471	-	7470A	PPB	15 Dec 2021 10:18:55	U	SMPL
M5074-07A - 1		17188	4.0826	-	7470A	PPB	15 Dec 2021 10:21:11	U	SMPL
M5072-03LX5 - 1		74	-0.0534	-	7470A	PPB	15 Dec 2021 10:23:28	U	SMPL
M5072-03A - 1		16890	4.0106	-	7470A	PPB	15 Dec 2021 10:25:45	U	SMPL
CCV22 - 1	CCV22	20813	4.9587	-	7470A	PPB	15 Dec 2021 10:28:02	U	SMPL
CCB22 - 1	CCB22	147	-0.0358	-	7470A	PPB	15 Dec 2021 10:30:18	U	SMPL