

LB 117802


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

53740



A= 0.0000e+000

B= 1.8485e-004

C= -4.2093e-002

Rho= 0.9997804

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%
0.0	0.000	0.008	0.008	271	0.000	271					-
0.2	0.200	0.190	-0.010	1258	0.0 %	1258					-5
2.5	2.500	2.441	-0.059	13434	0.0 %	13434					-2
5.0	5.000	5.028	0.028	27426	0.0 %	27426					1
7.5	7.500	7.641	0.141	41562	0.0 %	41562					2
10.0	10.000	9.892	-0.108	53740	0.0 %	53740					-1

LB117802 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 15 Dec 2021 10:57:14

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	271	-	0.0000	7471B	PPB	15 Dec 2021 11:01:11	S	Std
0.2 - 1	50.2	1258	-	0.2000	7471B	PPB	15 Dec 2021 11:03:28	S	Std
2.5 - 1	52.5	13434	-	2.5000	7471B	PPB	15 Dec 2021 11:05:45	S	Std
5.0 - 1	55	27426	-	5.0000	7471B	PPB	15 Dec 2021 11:08:02	S	Std
7.5 - 1	57.5	41562	-	7.5000	7471B	PPB	15 Dec 2021 11:10:19	S	Std
10.0 - 1	510	53740	-	10.0000	7471B	PPB	15 Dec 2021 11:15:11	S	Std
ICV13 - 1	ICV13	21312	3.8975	-	7471B	PPB	15 Dec 2021 11:19:39	U	SMPL
ICB13 - 1	ICB13	440	0.0392	-	7471B	PPB	15 Dec 2021 11:21:53	U	SMPL
CCV23 - 1	CCV23	27102	4.9678	-	7471B	PPB	15 Dec 2021 11:24:09	U	SMPL
CCB23 - 1	CCB23	249	0.0039	-	7471B	PPB	15 Dec 2021 11:26:24	U	SMPL
CRA - 1	CRA	1221	0.1836	-	7471B	PPB	15 Dec 2021 11:28:40	U	SMPL
HighStd - 1	HighStd	53262	9.8035	-	7471B	PPB	15 Dec 2021 11:30:56	U	SMPL
ChkStd - 1	ChkStd	38083	6.9977	-	7471B	PPB	15 Dec 2021 11:33:12	U	SMPL
PB141438BL - 1	PBS	136	-0.0170	-	7471B	PPB	15 Dec 2021 11:35:29	U	SMPL
PB141438BS - 1	LCSS	21473	3.9272	-	7471B	PPB	15 Dec 2021 11:37:47	U	SMPL
M5076-01 - 1	SB-11-C	347	0.0221	-	7471B	PPB	15 Dec 2021 11:40:04	U	SMPL
M5076-02 - 1	SB-11-G	295	0.0124	-	7471B	PPB	15 Dec 2021 11:42:21	U	SMPL
M5076-03 - 1	SB-15-C	1701	0.2723	-	7471B	PPB	15 Dec 2021 11:44:38	U	SMPL
M5076-04 - 1	SB-15-G	906	0.1254	-	7471B	PPB	15 Dec 2021 11:46:55	U	SMPL
M5076-04DUP - 1	SB-15-GDUP	830	0.1113	-	7471B	PPB	15 Dec 2021 11:49:10	U	SMPL
CCV24 - 1	CCV24	27240	4.9933	-	7471B	PPB	15 Dec 2021 11:51:25	U	SMPL
CCB24 - 1	CCB24	420	0.0355	-	7471B	PPB	15 Dec 2021 11:53:41	U	SMPL
M5076-05 - 1	M5076-04MS	21648	3.9596	-	7471B	PPB	15 Dec 2021 11:55:56	U	SMPL
M5076-06 - 1	M5076-04MSD	20928	3.8265	-	7471B	PPB	15 Dec 2021 11:58:12	U	SMPL
M5076-07 - 1	SB-18-C	1587	0.2513	-	7471B	PPB	15 Dec 2021 12:00:27	U	SMPL
M5076-08 - 1	SB-18-G	1610	0.2555	-	7471B	PPB	15 Dec 2021 12:02:42	U	SMPL
M5076-09 - 1	SB-19-C	2337	0.3899	-	7471B	PPB	15 Dec 2021 12:04:58	U	SMPL
M5076-10 - 1	SB-19-G	2743	0.4650	-	7471B	PPB	15 Dec 2021 12:07:15	U	SMPL
M5076-11 - 1	SB-9-SB-22-C	3090	0.5291	-	7471B	PPB	15 Dec 2021 12:09:31	U	SMPL
M5076-12 - 1	SB-9-SB-22-G	2750	0.4663	-	7471B	PPB	15 Dec 2021 12:11:48	U	SMPL
M5076-13 - 1	SB-20-C	719	0.0908	-	7471B	PPB	15 Dec 2021 12:14:05	U	SMPL
M5076-14 - 1	SB-20-G	318	0.0167	-	7471B	PPB	15 Dec 2021 12:16:20	U	SMPL
CCV25 - 1	CCV25	27279	5.0005	-	7471B	PPB	15 Dec 2021 12:18:36	U	SMPL
CCB25 - 1	CCB25	434	0.0381	-	7471B	PPB	15 Dec 2021 12:20:51	U	SMPL
PB141439BL - 1	PBS	136	-0.0170	-	7471B	PPB	15 Dec 2021 12:23:07	U	SMPL
PB141439BS - 1	LCSS	20893	3.8200	-	7471B	PPB	15 Dec 2021 12:25:23	U	SMPL
M5064-01 - 1	SASP2-121-1-(0-5)	2143	0.3540	-	7471B	PPB	15 Dec 2021 12:27:39	U	SMPL
M5064-01DUP - 1	SASP2-121-1-(0-5)DUP	2238	0.3716	-	7471B	PPB	15 Dec 2021 12:29:54	U	SMPL
M5064-01MS - 1	SASP2-121-1-(0-5)MS	22670	4.1485	-	7471B	PPB	15 Dec 2021 12:32:10	U	SMPL
M5064-01MSD - 1	SASP2-121-1-(0-5)MSD	22939	4.1982	-	7471B	PPB	15 Dec 2021 12:34:28	U	SMPL
M5064-02 - 1	SASP2-121-1-(5-10)	497	0.0498	-	7471B	PPB	15 Dec 2021 12:36:44	U	SMPL
M5064-03 - 1	SASP2-121-1-(10-17)	853	0.1156	-	7471B	PPB	15 Dec 2021 12:39:01	U	SMPL
M5064-04 - 1	SASP2-121-1-(17-32)	535	0.0568	-	7471B	PPB	15 Dec 2021 12:41:18	U	SMPL
M5068-01 - 1	PL-01-121321	1026	0.1476	-	7471B	PPB	15 Dec 2021 12:43:34	U	SMPL
CCV26 - 1	CCV26	27487	5.0390	-	7471B	PPB	15 Dec 2021 12:45:50	U	SMPL
CCB26 - 1	CCB26	324	0.0178	-	7471B	PPB	15 Dec 2021 12:48:05	U	SMPL
M5073-01 - 1	EO-02-121321	4481	0.7862	-	7471B	PPB	15 Dec 2021 12:50:21	U	SMPL
M5073-04 - 1	EO-03-121321	3076	0.5265	-	7471B	PPB	15 Dec 2021 12:52:36	U	SMPL
M5076-04LX5 - 1		184	-0.0081	-	7471B	PPB	15 Dec 2021 12:59:15	U	SMPL
M5076-04A - 1		21678	3.9651	-	7471B	PPB	15 Dec 2021 13:01:31	U	SMPL
M5064-01LX5 - 1		392	0.0304	-	7471B	PPB	15 Dec 2021 13:03:47	U	SMPL
M5064-01A - 1		24652	4.5149	-	7471B	PPB	15 Dec 2021 13:06:03	U	SMPL
CCV27 - 1	CCV27	27802	5.0972	-	7471B	PPB	15 Dec 2021 13:08:20	U	SMPL
CCB27 - 1	CCB27	90	-0.0255	-	7471B	PPB	15 Dec 2021 13:10:37	U	SMPL