



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 12/16/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 14:40
In Date: 12/15/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 12/16/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117812

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M5101-01	BGL22	1	1.18	8.53	9.71	9.2	94.0	
M5101-02	BGL23	2	1.14	8.62	9.76	8.54	85.8	
M5101-03	BGL24	3	1.17	8.69	9.86	8.3	82.0	
M5101-04	BGL25	4	1.11	8.84	9.95	7.82	75.9	
M5101-05	BGL26	5	1.18	8.79	9.97	9.08	89.9	
M5101-06	BGL27	6	1.13	8.75	9.88	7.31	70.6	
M5101-07	BGL28	7	1.15	8.66	9.81	7.35	71.6	
M5101-08	BGL31	8	1.18	8.48	9.66	8.88	90.8	
M5101-09	BGL32	9	1.13	8.58	9.71	8.9	90.6	
M5101-10	BGL33	10	1.19	8.78	9.97	8.38	81.9	
M5101-11	BGL35	11	1.19	8.73	9.92	7.79	75.6	
M5101-12	BGL36	12	1.15	8.73	9.88	8.92	89.0	
M5101-13	BGL37	13	1.19	8.79	9.98	8.73	85.8	
M5101-14	BGL38	14	1.19	8.58	9.77	7.45	73.0	
M5101-15	BGL39	15	1.15	8.82	9.97	7.64	73.6	
M5101-17	BGL41	16	1.15	8.80	9.95	9.44	94.2	
M5101-18	BGL42	17	1.19	8.63	9.82	9.07	91.3	
M5101-19	BGL43	18	1.18	8.51	9.69	8.37	84.5	
M5101-20	BGL44	19	1.17	8.59	9.76	8.05	80.1	
M5101-21	VHBLK001	20	1.00	1.00	2.00	2.00	100.0	VHBLK

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$