

LB117819


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

56088



A= 0.0000e+000

B= 1.7817e-004

C= -1.5027e-002

Rho= 0.9999858

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.006	0.006	117	0.000	117					1
0.2	0.200	0.192	-0.008	1162	0.0 %	1162					-4
2.5	2.500	2.500	-0.000	14114	0.0 %	14114					0
5.0	5.000	4.986	-0.014	28067	0.0 %	28067					0
7.5	7.500	7.539	0.039	42399	0.0 %	42399					1
10.0	10.000	9.978	-0.022	56088	0.0 %	56088					0

LB117819 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 15 Dec 2021 15:19:23

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	117	-	0.00007471B	PPB		15 Dec 2021 15:21:15	S	Std
0.2 - 1	50.2	1162	-	0.20007471B	PPB		15 Dec 2021 15:23:32	S	Std
2.5 - 1	52.5	14114	-	2.50007471B	PPB		15 Dec 2021 15:25:49	S	Std
5.0 - 1	55	28067	-	5.00007471B	PPB		15 Dec 2021 15:28:06	S	Std
7.5 - 1	57.5	42399	-	7.50007471B	PPB		15 Dec 2021 15:30:22	S	Std
10.0 - 1	510	56088	-	10.00007471B	PPB		15 Dec 2021 15:32:39	S	Std
ICV14 - 1	ICV14	21354	3.7895	-7471B	PPB		15 Dec 2021 15:35:19	U	SMPL
ICB14 - 1	ICB14	50	-0.0061	-7471B	PPB		15 Dec 2021 15:37:34	U	SMPL
CCV28 - 1	CCV28	26435	4.6948	-7471B	PPB		15 Dec 2021 15:39:49	U	SMPL
CCB28 - 1	CCB28	53	-0.0056	-7471B	PPB		15 Dec 2021 15:42:04	U	SMPL
CRA - 1	CRA	1085	0.1783	-7471B	PPB		15 Dec 2021 15:44:20	U	SMPL
HighStd - 1	HighStd	53915	9.5908	-7471B	PPB		15 Dec 2021 15:46:34	U	SMPL
ChkStd - 1	ChkStd	37706	6.7029	-7471B	PPB		15 Dec 2021 15:48:50	U	SMPL
PB141462BL - 1	PBS	-38	-0.0218	-7471B	PPB		15 Dec 2021 15:51:07	U	SMPL
PB141462BS - 1	LCSS	20389	3.6176	-7471B	PPB		15 Dec 2021 15:53:24	U	SMPL
M5078-01 - 1	340	1508	0.2536	-7471B	PPB		15 Dec 2021 15:55:40	U	SMPL
M5078-03 - 1	338	1822	0.3096	-7471B	PPB		15 Dec 2021 15:57:57	U	SMPL
M5079-01 - 1	VNJ-229	1385	0.2317	-7471B	PPB		15 Dec 2021 16:00:14	U	SMPL
M5079-04 - 1	RBR-21141	4509	0.7883	-7471B	PPB		15 Dec 2021 16:02:32	U	SMPL
M5081-01 - 1	CONCRETE	7927	1.3973	-7471B	PPB		15 Dec 2021 16:04:47	U	SMPL
CCV29 - 1	CCV29	27445	4.8747	-7471B	PPB		15 Dec 2021 16:07:03	U	SMPL
CCB29 - 1	CCB29	57	-0.0049	-7471B	PPB		15 Dec 2021 16:09:18	U	SMPL
M5083-02 - 1	BB1	13702	2.4262	-7471B	PPB		15 Dec 2021 16:11:34	U	SMPL
M5083-05 - 1	BB2	12603	2.2304	-7471B	PPB		15 Dec 2021 16:13:50	U	SMPL
M5083-08 - 1	BB3	12131	2.1463	-7471B	PPB		15 Dec 2021 16:16:05	U	SMPL
M5083-11 - 1	BB4	4415	0.7716	-7471B	PPB		15 Dec 2021 16:18:21	U	SMPL
M5083-14 - 1	BB5	4567	0.7987	-7471B	PPB		15 Dec 2021 16:20:37	U	SMPL
M5083-17 - 1	BB6	43057	7.6563	-7471B	PPB		15 Dec 2021 16:22:53	U	SMPL
M5083-17DUP - 1	BB6DUP	48963	8.7085	-7471B	PPB		15 Dec 2021 16:25:10	U	SMPL
M5083-17MS - 1	BB6MS	68375	12.1671	-7471B	PPB		15 Dec 2021 16:30:18	U	SMPL
M5083-17MSD - 1	BB6MSD	71352	12.6975	-7471B	PPB		15 Dec 2021 16:35:15	U	SMPL
M5086-01 - 1	14ST-SP-121421	997	0.1626	-7471B	PPB		15 Dec 2021 16:37:30	U	SMPL
CCV30 - 1	CCV30	26800	4.7598	-7471B	PPB		15 Dec 2021 16:39:47	U	SMPL
CCB30 - 1	CCB30	68	-0.0029	-7471B	PPB		15 Dec 2021 16:42:03	U	SMPL
M5087-01 - 1	OR-02-121421	3023	0.5236	-7471B	PPB		15 Dec 2021 16:44:19	U	SMPL
PB141463BL - 1	PBS	41	-0.0077	-7471B	PPB		15 Dec 2021 16:46:35	U	SMPL
PB141463BS - 1	LCSS	21196	3.7614	-7471B	PPB		15 Dec 2021 16:48:50	U	SMPL
M5088-01 - 1	IB-1-0.5	123	0.0069	-7471B	PPB		15 Dec 2021 16:51:06	U	SMPL
M5088-02 - 1	IB-1-1.5	1353	0.2260	-7471B	PPB		15 Dec 2021 16:53:22	U	SMPL
M5088-03 - 1	IB-2-0.5	161	0.0137	-7471B	PPB		15 Dec 2021 16:55:38	U	SMPL
M5088-04 - 1	IB-2-1.5	189	0.0186	-7471B	PPB		15 Dec 2021 16:57:54	U	SMPL
M5088-05 - 1	IB-3-0.5	264	0.0320	-7471B	PPB		15 Dec 2021 17:00:11	U	SMPL
M5088-06 - 1	IB-3-1.5	272	0.0334	-7471B	PPB		15 Dec 2021 17:02:28	U	SMPL
M5088-07 - 1	IB-4-0.5	262	0.0317	-7471B	PPB		15 Dec 2021 17:04:43	U	SMPL
CCV31 - 1	CCV31	27192	4.8297	-7471B	PPB		15 Dec 2021 17:06:59	U	SMPL
CCB31 - 1	CCB31	19	-0.0116	-7471B	PPB		15 Dec 2021 17:09:14	U	SMPL
M5088-08 - 1	IB-4-1.5	255	0.0304	-7471B	PPB		15 Dec 2021 17:11:31	U	SMPL
M5088-08DUP - 1	IB-4-1.5DUP	259	0.0311	-7471B	PPB		15 Dec 2021 17:13:46	U	SMPL
M5088-08MS - 1	IB-4-1.5MS	20826	3.6955	-7471B	PPB		15 Dec 2021 17:16:02	U	SMPL
M5088-08MSD - 1	IB-4-1.5MSD	20869	3.7031	-7471B	PPB		15 Dec 2021 17:18:17	U	SMPL
M5089-02 - 1	MS-1	1332	0.2223	-7471B	PPB		15 Dec 2021 17:20:33	U	SMPL
M5089-06 - 1	MS-2	2762	0.4771	-7471B	PPB		15 Dec 2021 17:22:50	U	SMPL
M5089-10 - 1	MS-3	3551	0.6176	-7471B	PPB		15 Dec 2021 17:25:07	U	SMPL
M5089-14 - 1	MS-4	10569	1.8680	-7471B	PPB		15 Dec 2021 17:27:23	U	SMPL
CCV32 - 1	CCV32	27499	4.8844	-7471B	PPB		15 Dec 2021 17:29:40	U	SMPL
CCB32 - 1	CCB32	14	-0.0125	-7471B	PPB		15 Dec 2021 17:31:56	U	SMPL
M5083-17LX5 - 1		8323	1.4679	-7471B	PPB		15 Dec 2021 17:34:12	U	SMPL
M5083-17A - 1		71961	12.8060	-7471B	PPB		15 Dec 2021 17:36:28	U	SMPL
M5088-08LX5 - 1		-84	-0.0300	-7471B	PPB		15 Dec 2021 17:38:43	U	SMPL
M5088-08A - 1		20915	3.7113	-7471B	PPB		15 Dec 2021 17:41:01	U	SMPL
M5083-17DLX2 - 1		21265	3.7737	-7471B	PPB		15 Dec 2021 17:43:16	U	SMPL
M5083-17DUPDLX2 - 1		25188	4.4726	-7471B	PPB		15 Dec 2021 17:45:32	U	SMPL
M5083-17MSDLX2 - 1		35151	6.2477	-7471B	PPB		15 Dec 2021 17:47:48	U	SMPL
M5083-17MSDDLX2 - 1		34425	6.1183	-7471B	PPB		15 Dec 2021 17:50:05	U	SMPL
M5083-17LDLX10 - 1		2783	0.4808	-7471B	PPB		15 Dec 2021 17:52:22	U	SMPL
CCV33 - 1	CCV33	27272	4.8439	-7471B	PPB		15 Dec 2021 17:56:58	U	SMPL
CCB33 - 1	CCB33	41	-0.0077	-7471B	PPB		15 Dec 2021 17:59:14	U	SMPL