



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB117824

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/15/2021	16:00
2	CAL2	1	Water	NA	NA	20.3	7.00	12/15/2021	16:01
3	CAL3	1	Water	NA	NA	20.3	10.03	12/15/2021	16:02
4	ICV	1	Water	NA	NA	20.4	7.02	12/15/2021	16:05
5	CCV1	1	Water	NA	NA	20.4	2.01	12/15/2021	16:10
6	M5051-05	1	Solid	20.03	20	20.4	8.53	12/15/2021	16:11
7	M5051-06	1	Solid	20.04	20	20.3	9.11	12/15/2021	16:15
8	M5051-07	1	Solid	20.03	20	20.5	8.80	12/15/2021	16:20
9	M5051-08	1	Solid	20.04	20	20.9	7.61	12/15/2021	16:22
10	M5097-02	1	Solid	20.03	20	20.6	6.97	12/15/2021	16:25
11	M5097-02DUP	1	Solid	20.04	20	20.7	6.98	12/15/2021	16:26
12	CCV2	1	Water	NA	NA	20.4	12.03	12/15/2021	16:30

117824

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M5097 WorkList ID : 155219 Department : Wet-Chemistry Date : 12-15-2021 15:00:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/10/2021	Solid	M5051-05	Corrosivity	Cool 4 deg C	ENTA05	P-1	P-1	12/10/2021	9045D
12/10/2021	Solid	M5051-06	Corrosivity	Cool 4 deg C	ENTA05	P-2	P-2	12/10/2021	9045D
12/10/2021	Solid	M5051-07	Corrosivity	Cool 4 deg C	ENTA05	P-3	P-3	12/10/2021	9045D
12/10/2021	Solid	M5051-08	Corrosivity	Cool 4 deg C	ENTA05	A5	A5	12/10/2021	9045D
12/15/2021	Solid	M5097-02	Corrosivity	Cool 4 deg C	PSEG03	H32	ETGI-371	12/15/2021	9045D

Date/Time 12-15-21 15:10
Raw Sample Received by: JP (WJ)
Raw Sample Relinquished by: JP (WJ)

Date/Time 12-15-21 17:00
Raw Sample Received by: JP (WJ)
Raw Sample Relinquished by: JP (WJ)