


LB117828

SFAM01.1

INSTRUMENT ID: CVL

 μ Abs.:

43200



Linear

A= 0.0000e+000

B= 2.3138e-004 *slope*C= -1.4336e-002 *y-intercept*

Rho= 0.9999915

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>%D</i>
0.00	0.000	0.003	0.003	77	0.000	77					
0.05	0.050					205	0				
0.20	0.200	0.188	-0.012	873	0.0 %	873					
2.50	2.500	2.491	-0.009	10827	0.0 %	10827					
5.00	5.000	5.026	0.026	21783	0.0 %	21783					
7.50	7.500	7.511	0.011	32525	0.0 %	32525					
10.0	10.000	9.981	-0.019	43200	0.0 %	43200					

%D
 1
 6
 0
 1
 0
 0
 0

LB117828
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	77	-		0 SFAM01.1	PPB	12/16/2021 10:36	Std
	0.2 S01	873	-		0.2 SFAM01.1	PPB	12/16/2021 10:38	Std
	2.5 S02	10827	-		2.5 SFAM01.1	PPB	12/16/2021 10:41	Std
	5 S03	21783	-		5 SFAM01.1	PPB	12/16/2021 10:43	Std
	7.5 S04	32525	-		7.5 SFAM01.1	PPB	12/16/2021 10:45	Std
	10 S05	43200	-		10 SFAM01.1	PPB	12/16/2021 10:48	Std
ICV097	ICV097	16930	3.9029 -		SFAM01.1	PPB	12/16/2021 10:50	SMPL
ICB097	ICB097	-13	-0.0173 -		SFAM01.1	PPB	12/16/2021 10:53	SMPL
CCV034	CCV034	22292	5.1435 -		SFAM01.1	PPB	12/16/2021 10:55	SMPL
CCB034	CCB034	-16	-0.018 -		SFAM01.1	PPB	12/16/2021 10:57	SMPL
PB141464BL	PBW464	51	-0.0025 -		SFAM01.1	PPB	12/16/2021 10:59	SMPL
M5000-18	A0BC9	479	0.0965 -		SFAM01.1	PPB	12/16/2021 11:02	SMPL
PB141465BL	PBW465	44	-0.0042 -		SFAM01.1	PPB	12/16/2021 11:04	SMPL
M5001-13	A0BD0	469	0.0942 -		SFAM01.1	PPB	12/16/2021 11:06	SMPL
PB141466BL	PBW466	32	-0.0069 -		SFAM01.1	PPB	12/16/2021 11:08	SMPL
M5040-01	C00V3	33	-0.0067 -		SFAM01.1	PPB	12/16/2021 11:11	SMPL
M5040-02	C00V5	39	-0.0053 -		SFAM01.1	PPB	12/16/2021 11:13	SMPL
M5040-03	C00V7	37	-0.0058 -		SFAM01.1	PPB	12/16/2021 11:15	SMPL
M5040-04	C00V9	40	-0.0051 -		SFAM01.1	PPB	12/16/2021 11:17	SMPL
M5040-05	C00W1	52	-0.0023 -		SFAM01.1	PPB	12/16/2021 11:20	SMPL
M5040-06	C00W3	23	-0.009 -		SFAM01.1	PPB	12/16/2021 11:22	SMPL
M5040-07	C00W5	10	-0.012 -		SFAM01.1	PPB	12/16/2021 11:24	SMPL
M5040-08	C00W7	12	-0.0116 -		SFAM01.1	PPB	12/16/2021 11:26	SMPL
M5040-09	C00W9	14	-0.0111 -		SFAM01.1	PPB	12/16/2021 11:29	SMPL
M5040-10	C00W9D	35	-0.0062 -		SFAM01.1	PPB	12/16/2021 11:31	SMPL
M5040-11	C00W9S	3506	0.7969 -		SFAM01.1	PPB	12/16/2021 11:33	SMPL
M5040-12	C00X2	19	-0.0099 -		SFAM01.1	PPB	12/16/2021 11:36	SMPL
M5040-13	C00X4	26	-0.0083 -		SFAM01.1	PPB	12/16/2021 11:38	SMPL
M5040-14	C00X6	17	-0.0104 -		SFAM01.1	PPB	12/16/2021 11:40	SMPL
M5040-15	C00X8	28	-0.0079 -		SFAM01.1	PPB	12/16/2021 11:42	SMPL
CCV035	CCV035	22425	5.1743 -		SFAM01.1	PPB	12/16/2021 11:45	SMPL
CCB035	CCB035	-27	-0.0206 -		SFAM01.1	PPB	12/16/2021 11:47	SMPL
M5040-16	C00Y0	19	-0.0099 -		SFAM01.1	PPB	12/16/2021 11:49	SMPL
M5040-17	C00Y2	16	-0.0106 -		SFAM01.1	PPB	12/16/2021 11:51	SMPL
M5040-18	C00Y4	14	-0.0111 -		SFAM01.1	PPB	12/16/2021 11:54	SMPL
M5040-19	C00Y6	27	-0.0081 -		SFAM01.1	PPB	12/16/2021 11:56	SMPL
M5040-20	C00Y8	1	-0.0141 -		SFAM01.1	PPB	12/16/2021 11:58	SMPL
M5040-21	C00Z0	83	0.0049 -		SFAM01.1	PPB	12/16/2021 12:00	SMPL
M5040-22	C00Z2	14	-0.0111 -		SFAM01.1	PPB	12/16/2021 12:03	SMPL
PB141467BL	PBW467	46	-0.0037 -		SFAM01.1	PPB	12/16/2021 12:05	SMPL
M5041-01	C00Z4	0	-0.0143 -		SFAM01.1	PPB	12/16/2021 12:07	SMPL
PB141468BL	PBW468	32	-0.0069 -		SFAM01.1	PPB	12/16/2021 12:10	SMPL
M5042-01	C00V6	4	-0.0134 -		SFAM01.1	PPB	12/16/2021 12:12	SMPL
M5042-02	C00V8	17	-0.0104 -		SFAM01.1	PPB	12/16/2021 12:14	SMPL
M5042-03	C00W0	35	-0.0062 -		SFAM01.1	PPB	12/16/2021 12:16	SMPL
M5042-04	C00W2	18	-0.0102 -		SFAM01.1	PPB	12/16/2021 12:19	SMPL

INSTRUMENT ID : CV1

M5042-05	C00W4	16	-0.0106 -	SFAM01.1	PPB	12/16/2021 12:21	SMPL
M5042-06	C00W6	-3	-0.015 -	SFAM01.1	PPB	12/16/2021 12:23	SMPL
M5042-07	C00W8	3	-0.0136 -	SFAM01.1	PPB	12/16/2021 12:25	SMPL
M5042-08	C00X0	32	-0.0069 -	SFAM01.1	PPB	12/16/2021 12:28	SMPL
M5042-09	C00X0D	16	-0.0106 -	SFAM01.1	PPB	12/16/2021 12:30	SMPL
M5042-10	C00X0S	3461	0.7865 -	SFAM01.1	PPB	12/16/2021 12:32	SMPL
CCV036	CCV036	21829	5.0364 -	SFAM01.1	PPB	12/16/2021 12:35	SMPL
CCB036	CCB036	-25	-0.0201 -	SFAM01.1	PPB	12/16/2021 12:37	SMPL
M5042-11	C00X3	17	-0.0104 -	SFAM01.1	PPB	12/16/2021 12:39	SMPL
M5042-12	C00X5	31	-0.0072 -	SFAM01.1	PPB	12/16/2021 12:41	SMPL
M5042-13	C00X7	18	-0.0102 -	SFAM01.1	PPB	12/16/2021 12:44	SMPL
M5042-14	C00X9	23	-0.009 -	SFAM01.1	PPB	12/16/2021 12:46	SMPL
M5042-15	C00Y1	34	-0.0065 -	SFAM01.1	PPB	12/16/2021 12:48	SMPL
M5042-16	C00Y3	8	-0.0125 -	SFAM01.1	PPB	12/16/2021 12:50	SMPL
M5042-17	C00Y5	25	-0.0086 -	SFAM01.1	PPB	12/16/2021 12:53	SMPL
M5042-18	C00Y7	33	-0.0067 -	SFAM01.1	PPB	12/16/2021 12:55	SMPL
M5042-19	C00Y9	55	-0.0016 -	SFAM01.1	PPB	12/16/2021 12:57	SMPL
M5042-20	C00Z1	31	-0.0072 -	SFAM01.1	PPB	12/16/2021 13:00	SMPL
M5042-21	C00Z3	18	-0.0102 -	SFAM01.1	PPB	12/16/2021 13:02	SMPL
M5042-22	C00Z5	7	-0.0127 -	SFAM01.1	PPB	12/16/2021 13:04	SMPL
PB141469BL	PBW469	118	0.013 -	SFAM01.1	PPB	12/16/2021 13:06	SMPL
M5046-01	MH1D92	222	0.037 -	SFAM01.1	PPB	12/16/2021 13:09	SMPL
M5046-02	MH1D94	268	0.0477 -	SFAM01.1	PPB	12/16/2021 13:11	SMPL
M5046-03	MH1D96	255	0.0447 -	SFAM01.1	PPB	12/16/2021 13:13	SMPL
M5046-04	MH1D98	275	0.0493 -	SFAM01.1	PPB	12/16/2021 13:15	SMPL
M5046-05	MH1E00	157	0.022 -	SFAM01.1	PPB	12/16/2021 13:18	SMPL
M5046-06	MH1E02	125	0.0146 -	SFAM01.1	PPB	12/16/2021 13:20	SMPL
M5046-07	MH1E04	154	0.0213 -	SFAM01.1	PPB	12/16/2021 13:22	SMPL
CCV037	CCV037	21938	5.0616 -	SFAM01.1	PPB	12/16/2021 13:24	SMPL
CCB037	CCB037	-28	-0.0208 -	SFAM01.1	PPB	12/16/2021 13:27	SMPL
M5046-08	MH1E06	100	0.0088 -	SFAM01.1	PPB	12/16/2021 13:29	SMPL
M5046-09	MH1E08	322	0.0602 -	SFAM01.1	PPB	12/16/2021 13:31	SMPL
M5046-10	MH1E10	12059	2.7758 -	SFAM01.1	PPB	12/16/2021 13:34	SMPL
M5046-11	MH1E12	520	0.106 -	SFAM01.1	PPB	12/16/2021 13:36	SMPL
M5046-12	MH1E14	388	0.0754 -	SFAM01.1	PPB	12/16/2021 13:38	SMPL
M5046-13	MH1E16	423	0.0835 -	SFAM01.1	PPB	12/16/2021 13:40	SMPL
M5046-14	MH1E18	150	0.0204 -	SFAM01.1	PPB	12/16/2021 13:43	SMPL
M5046-15	MH1E20	617	0.1284 -	SFAM01.1	PPB	12/16/2021 13:45	SMPL
M5046-16	MH1E22	628	0.131 -	SFAM01.1	PPB	12/16/2021 13:47	SMPL
M5046-17	MH1E24	765	0.1627 -	SFAM01.1	PPB	12/16/2021 13:49	SMPL
M5046-18	MH1E26	862	0.1851 -	SFAM01.1	PPB	12/16/2021 13:52	SMPL
M5046-19	MH1E28	1011	0.2196 -	SFAM01.1	PPB	12/16/2021 13:54	SMPL
M5046-20	MH1E30	481	0.097 -	SFAM01.1	PPB	12/16/2021 13:56	SMPL
M5046-21	MH1E30D	481	0.097 -	SFAM01.1	PPB	12/16/2021 13:59	SMPL
M5046-22	MH1E30S	4425	1.0095 -	SFAM01.1	PPB	12/16/2021 14:01	SMPL
CCV038	CCV038	22506	5.193 -	SFAM01.1	PPB	12/16/2021 14:03	SMPL
CCB038	CCB038	-6	-0.0157 -	SFAM01.1	PPB	12/16/2021 14:05	SMPL