



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117831

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/16/2021	12:00
2	CAL2	1	Water	NA	NA	20.3	7.00	12/16/2021	12:01
3	CAL3	1	Water	NA	NA	20.3	10.03	12/16/2021	12:04
4	ICV	1	Water	NA	NA	20.4	6.99	12/16/2021	12:05
5	CCV1	1	Water	NA	NA	20.3	2.01	12/16/2021	12:06
6	M5120-01	1	Water	NA	NA	20.7	5.46	12/16/2021	12:10
7	M5120-01DUP	1	Water	NA	NA	20.8	5.47	12/16/2021	12:11
8	CCV2	1	Water	NA	NA	20.4	12.03	12/16/2021	12:15

1311783)

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M5120 WorkList ID : 155247 Department : Wet-Chemistry Date : 12-16-2021 11:41:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/14/2021	Water	M5120-01	pH	Cool 4 deg C	STER08	H41	POOL-WATER	12/14/2021	9040C

Date/Time 12.16.2021 11:45
Raw Sample Received by: JO (WC)
Raw Sample Relinquished by: NR SM

Date/Time 12.16.2021 12:35
Raw Sample Received by: CP SM
Raw Sample Relinquished by: JO (WC)