



Analytical Summary Report

Analysis Method: SM2510 B

ANALYST: Iwona

Parameter: Conductance

SUPERVISOR REVIEW BY: apatel

Run Number: LB117834

Meter ID: WC CONDUCTANCE METER-1

Thermometer ID: WC PH METER-1

Reagent/Standard	Lot/Log #
conductivity standard	WP92563

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB117834BL		1	21.5	0.86	0.92	12/16/2021	13:07
2	LB117834BS	1413	1	21.8	1326.00	1412.32	12/16/2021	13:09
3	M5120-01		1	18.8	2.12	2.40	12/16/2021	13:11
4	M5120-01DUP		1	18.9	2.13	2.41	12/16/2021	13:11

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$