



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB117838

Slope : 98.4

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/16/2021	15:15
2	CAL2	1	Water	NA	NA	20.3	7.00	12/16/2021	15:16
3	CAL3	1	Water	NA	NA	20.3	10.03	12/16/2021	15:17
4	ICV	1	Water	NA	NA	20.4	7.02	12/16/2021	15:19
5	CCV1	1	Water	NA	NA	20.4	2.01	12/16/2021	15:20
6	M5114-02	1	Solid	20.03	20	20.3	6.36	12/16/2021	15:22
7	M5124-02	1	Solid	20.04	20	20.3	8.73	12/16/2021	15:25
8	M5129-01	1	Solid	20.03	20	20.2	11.07	12/16/2021	15:30
9	M5129-03	1	Solid	20.04	20	20.3	11.34	12/16/2021	15:31
10	M5129-03DUP	1	Solid	20.03	20	20.2	11.35	12/16/2021	15:33
11	CCV2	1	Water	NA	NA	20.4	12.03	12/16/2021	15:37

W117838

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M5129 WorkList ID : 155266 Department : Wet-Chemistry Date : 12-16-2021 15:06:57

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/16/2021	Solid	M5114-02	Corrosivity	Cool 4 deg C	PSEG03	H32	HS-OILY-DEBRIS	12/16/2021	9045D
12/16/2021	Solid	M5124-02	Corrosivity	Cool 4 deg C	PSEG03	H41	F2337-STOPJOINT-TP	12/16/2021	9045D
12/16/2021	Solid	M5129-01	Corrosivity	Cool 4 deg C	PSEG03	H41	48-CL-1	12/16/2021	9045D
12/16/2021	Solid	M5129-03	Corrosivity	Cool 4 deg C	PSEG03	H41	48-CL-2	12/16/2021	9045D

Date/Time 12-16-21 15:10
 Raw Sample Received by: JP (QC)
 Raw Sample Relinquished by: JP (QC)

Date/Time 12-16-21 14:00
 Raw Sample Received by: JP (QC)
 Raw Sample Relinquished by: JP (QC)