



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB117840

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/16/2021	15:15
2	CAL2	1	Water	NA	NA	20.3	7.00	12/16/2021	15:16
3	CAL3	1	Water	NA	NA	20.3	10.03	12/16/2021	15:17
4	ICV	1	Water	NA	NA	20.4	7.02	12/16/2021	15:19
5	CCV1	1	Water	NA	NA	20.3	2.01	12/16/2021	16:25
6	M5114-03	1	Water	NA	NA	20.1	8.41	12/16/2021	16:30
7	M5114-03DUP	1	Water	NA	NA	20.2	8.42	12/16/2021	16:31
8	CCV2	1	Water	NA	NA	20.4	12.03	12/16/2021	16:35

12117840

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M5114 V WorkList ID : 155268 Department : Wet-Chemistry Date : 12-16-2021 16:14:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/16/2021	Water	M5114-03	Corrosivity	Cool 4 deg C	PSEG03	H32	HS-OILY-WATER	12/16/2021	9040C

Date/Time 12-16-21 16:20
 Raw Sample Received by: JN (COC)
 Raw Sample Relinquished by: JN (COC)

Date/Time 12-16-21 17:00
 Raw Sample Received by: JN (COC)
 Raw Sample Relinquished by: JN (COC)