


LB 117841

7471B

INSTRUMENT ID: CY2



Linear

A= 0.0000e+000

B= 1.7774e-004

C= 1.5444e-002

Rho= 0.9997919

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.033	0.033	101	0.000	101					-
0.2	0.200	0.206	0.006	1074	0.0 %	1074					3
2.5	2.500	2.391	-0.109	13368	0.0 %	13368					-4
5.0	5.000	5.032	0.032	28225	0.0 %	28225					1
7.5	7.500	7.617	0.117	42767	0.0 %	42767					2
10.0	10.000	9.920	-0.080	55726	0.0 %	55726					-1

LB117841 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 17 Dec 2021 09:17:51

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	101	-	0.00007471B	PPB	17 Dec 2021 09:18:10	S	Std
0.2 - 1	50.2	1074	-	0.20007471B	PPB	17 Dec 2021 09:20:27	S	Std
2.5 - 1	52.5	13368	-	2.50007471B	PPB	17 Dec 2021 09:22:45	S	Std
5.0 - 1	55	28225	-	5.00007471B	PPB	17 Dec 2021 09:25:02	S	Std
7.5 - 1	57.5	42767	-	7.50007471B	PPB	17 Dec 2021 09:27:18	S	Std
10.0 - 1	510	55726	-	10.00007471B	PPB	17 Dec 2021 09:32:07	S	Std
ICV15 - 1	ICV15	22580	4.0288	-7471B	PPB	17 Dec 2021 09:34:49	U	SMPL
ICB15 - 1	ICB15	38	0.0222	-7471B	PPB	17 Dec 2021 09:37:04	U	SMPL
CCV39 - 1	CCV39	27997	4.9916	-7471B	PPB	17 Dec 2021 09:39:20	U	SMPL
CCB39 - 1	CCB39	43	0.0231	-7471B	PPB	17 Dec 2021 09:41:35	U	SMPL
CRA - 1	CRA	1058	0.2035	-7471B	PPB	17 Dec 2021 09:43:50	U	SMPL
HighStd - 1	HighStd	55497	9.8793	-7471B	PPB	17 Dec 2021 09:46:07	U	SMPL
ChkStd - 1	ChkStd	40401	7.1962	-7471B	PPB	17 Dec 2021 09:48:22	U	SMPL
PB141502BL - 1	PBS	84	0.0304	-7471B	PPB	17 Dec 2021 09:50:39	U	SMPL
PB141502BS - 1	LCSS	22576	4.0280	-7471B	PPB	17 Dec 2021 09:55:31	U	SMPL
M5082-01 - 1	SASP2-121-1(35-52)	640	0.1292	-7471B	PPB	17 Dec 2021 09:57:48	U	SMPL
M5082-02 - 1	SASP2-121-1(55-62)	1323	0.2506	-7471B	PPB	17 Dec 2021 10:00:05	U	SMPL
M5082-03 - 1	SASP2-121-1(65-77.7)	389	0.0846	-7471B	PPB	17 Dec 2021 10:02:22	U	SMPL
M5097-01 - 1	ETGI-371	801	0.1578	-7471B	PPB	17 Dec 2021 10:04:40	U	SMPL
M5098-01 - 1	DRUM-1-2	304	0.0695	-7471B	PPB	17 Dec 2021 10:06:55	U	SMPL
CCV40 - 1	CCV40	27843	4.9642	-7471B	PPB	17 Dec 2021 10:09:10	U	SMPL
CCB40 - 1	CCB40	72	0.0282	-7471B	PPB	17 Dec 2021 10:11:26	U	SMPL
M5105-01 - 1	HD-01-121521	1781	0.3320	-7471B	PPB	17 Dec 2021 10:13:42	U	SMPL
M5106-01 - 1	VAULT-WASTE	222	0.0549	-7471B	PPB	17 Dec 2021 10:15:57	U	SMPL
M5108-01 - 1	MW4CD-1.5	371	0.0814	-7471B	PPB	17 Dec 2021 10:18:13	U	SMPL
M5108-02 - 1	MW4CD-5	2600	0.4776	-7471B	PPB	17 Dec 2021 10:20:29	U	SMPL
M5108-03 - 1	MW4CD-10	395	0.0856	-7471B	PPB	17 Dec 2021 10:22:44	U	SMPL
M5108-04 - 1	MW14-1.5	1505	0.2829	-7471B	PPB	17 Dec 2021 10:25:01	U	SMPL
M5108-05 - 1	MW14-5	1426	0.2689	-7471B	PPB	17 Dec 2021 10:27:17	U	SMPL
M5108-06 - 1	MW14-10	977	0.1891	-7471B	PPB	17 Dec 2021 10:29:34	U	SMPL
M5108-07 - 1	MW15-1.5	1784	0.3325	-7471B	PPB	17 Dec 2021 10:31:51	U	SMPL
M5108-08 - 1	MW15-5	987	0.1909	-7471B	PPB	17 Dec 2021 10:34:07	U	SMPL
CCV41 - 1	CCV41	27822	4.9604	-7471B	PPB	17 Dec 2021 10:36:23	U	SMPL
CCB41 - 1	CCB41	48	0.0240	-7471B	PPB	17 Dec 2021 10:38:38	U	SMPL
M5108-09 - 1	MW15-10	597	0.1216	-7471B	PPB	17 Dec 2021 10:40:54	U	SMPL
M5108-10 - 1	MW16-1.5	2657	0.4877	-7471B	PPB	17 Dec 2021 10:43:09	U	SMPL
M5108-11 - 1	MW16-5	701	0.1400	-7471B	PPB	17 Dec 2021 10:45:25	U	SMPL
M5108-12 - 1	MW16-10	815	0.1603	-7471B	PPB	17 Dec 2021 10:47:41	U	SMPL
M5108-12DUP - 1	MW16-10DUP	763	0.1511	-7471B	PPB	17 Dec 2021 10:49:56	U	SMPL
M5108-12MS - 1	MW16-10MS	22515	4.0172	-7471B	PPB	17 Dec 2021 10:54:48	U	SMPL
M5108-12MSD - 1	MW16-10MSD	22590	4.0305	-7471B	PPB	17 Dec 2021 10:59:37	U	SMPL
M5108-12LX5 - 1		485	0.1016	-7471B	PPB	17 Dec 2021 11:01:54	U	SMPL
M5108-12A - 1		22477	4.0104	-7471B	PPB	17 Dec 2021 11:04:10	U	SMPL
CCV42 - 1	CCV42	27992	4.9907	-7471B	PPB	17 Dec 2021 11:06:26	U	SMPL
CCB42 - 1	CCB42	56	0.0254	-7471B	PPB	17 Dec 2021 11:08:42	U	SMPL