



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB117843

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	

Intercept: 0.2889

Slope: 0.8914

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.152	-0.15		12/17/2021	09:30
2	CAL2	1	1	1.040	0.84	15.7	12/17/2021	09:33
3	CAL3	5	1	4.420	4.63	7.3	12/17/2021	09:36
4	CAL4	10	1	9.350	10.17	-1.6	12/17/2021	09:40
5	CAL5	20	1	17.200	18.97	5.1	12/17/2021	09:43
6	CAL6	40	1	36.500	40.62	-1.6	12/17/2021	09:46
7	CAL7	80	1	73.000	81.57	-2	12/17/2021	09:50
8	CAL8	200	1	178.000	199.36	0.3	12/17/2021	09:53



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB117843

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.210	10.008	12/17/2021	09:56
2	ICB		1	0.148	-0.158	12/17/2021	10:00
3	CCV1	10	1	9.340	10.154	12/17/2021	10:03
4	CCB1		1	0.174	-0.129	12/17/2021	10:05
5	LB117843BL		1	0.182	-0.120	12/17/2021	10:08
6	M4068-12		1	0.624	0.376	12/17/2021	10:12
7	M5128-01		1	0.325	0.040	12/17/2021	10:15
8	M5128-03		1	163.000	182.534	12/17/2021	10:17
9	M5128-05		1	0.265	-0.027	12/17/2021	10:19
10	M5128-07		1	138.000	154.489	12/17/2021	10:22
11	M5128-07DUP		1	136.000	152.245	12/17/2021	10:24
12	CCV2	10	1	8.740	9.481	12/17/2021	10:27
13	CCB2		1	0.154	-0.151	12/17/2021	10:30

WORKLIST(Hardcopy Internal Chain)

LB117843

WorkList Name : TURB-121621 WorkList ID : 155261 Department : Wet-Chemistry Date : 12-16-2021 14:45:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/17/2021	Water	M5128-01	Turbidity	Cool 4 deg C	LOCK01	H41	MW-1	12/15/2021	SM2130 B
12/17/2021	Water	M5128-03	Turbidity	Cool 4 deg C	LOCK01	H41	MW-2	12/15/2021	SM2130 B
12/17/2021	Water	M5128-05	Turbidity	Cool 4 deg C	LOCK01	H41	MW-3	12/15/2021	SM2130 B
12/17/2021	Water	M5128-07	Turbidity	Cool 4 deg C	LOCK01	H41	MW-4	12/15/2021	SM2130 B

Date/Time 12/17/21 8:15
 Raw Sample Received by: 12(10)
 Raw Sample Relinquished by: 12(10)

Date/Time 12/17/21 19:30
 Raw Sample Received by: 12(10)
 Raw Sample Relinquished by: 12(10)