



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB117844

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP94436
Turbidity Calibration std, 0NTU	WP94435
Turbidity Calibration std, 20NTU	WP94430
Turbidity Calibration std, 80NTU	WP94428
Turbidity Calibration std, 40NTU	WP94429
Turbidity Calibration std, 200NTU	WP94427
Turbidity Calibration std, 1NTU	WP94433
Turbidity Calibration std, 5NTU	WP94432
Turbidity Calibration - CCV std, 10 NTU	WP94431

Intercept: 0.2502

Slope: 0.8493

Regression: 0.9997

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.152	-0.12		12/17/2021	09:30
2	CAL2	1	1	1.043	0.93	6.7	12/17/2021	09:31
3	CAL3	5	1	4.400	4.89	2.3	12/17/2021	09:32
4	CAL4	10	1	9.350	10.71	-7.1	12/17/2021	09:33
5	CAL5	20	1	16.800	19.49	2.6	12/17/2021	09:34
6	CAL6	40	1	34.300	40.09	-0.2	12/17/2021	09:35



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB117844

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.210	10.550	12/17/2021	09:36
2	ICB		1	0.148	-0.120	12/17/2021	09:40
3	CCV1	10	1	9.340	10.703	12/17/2021	09:41
4	CCB1		1	0.174	-0.090	12/17/2021	09:42
5	LB117844BL		1	0.182	-0.080	12/17/2021	09:43
6	M4068-09		1	0.624	0.440	12/17/2021	09:44
7	CCV2	10	1	8.740	9.996	12/17/2021	09:50
8	CCB2		1	0.154	-0.113	12/17/2021	09:51