



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB117847

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	12/17/2021	14:15
2	M5114-03		1	10	100.00	>212.0	>212.0	12/17/2021	14:40
3	M5135-01		1	12	100.00	>212.0	>212.0	12/17/2021	15:05
4	M5135-01DUP		1	12	100.00	>212.0	>212.0	12/17/2021	15:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB117847

WorkList Name : fp121721

WorkList ID : 155288

Department : Wet-Chemistry

Date : 12-17-2021 10:11:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/23/2021	Water	M5114-03	Flash Point	Cool 4 deg C	PSEG03	H32	HS-OILY-WATER	12/16/2021	1010B
12/23/2021	Water	M5135-01	Flash Point	Cool 4 deg C	PSEG03	H41	34009	12/16/2021	1010B

Date/Time

12/17/21 13:00

Raw Sample Received by:

KS (WC)

Raw Sample Relinquished by:

RJ C FZA- (WC)

Date/Time

12/17/21 15:40

Raw Sample Received by:

RJ C FZA- (WC)

Raw Sample Relinquished by:

KS (WC)