

LB117842


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

39783



A= 0.0000e+000

B= 2.5407e-004

C= -1.7038e-001

Rho= 0.9999135

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	90)
0.0	0.000	-0.034	-0.034	538	0.000	538					4
0.2	0.200	0.208	0.008	1490	0.0 %	1490					-1
2.5	2.500	2.469	-0.031	10389	0.0 %	10389					2
5.0	5.000	5.082	0.082	20672	0.0 %	20672					0
7.5	7.500	7.537	0.037	30335	0.0 %	30335					-1
10.0	10.000	9.937	-0.063	39783	0.0 %	39783					

LB117848 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 17 Dec 2021 12:08:46

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	30	538	-	0.0000	7470A	PPB	17 Dec 2021 12:15:04	S	Std
0.2 - 1	582	1490	-	0.2000	7470A	PPB	17 Dec 2021 12:17:21	S	Std
2.5 - 1	52.5	10389	-	2.5000	7470A	PPB	17 Dec 2021 12:19:38	S	Std
5.0 - 1	55	20672	-	5.0000	7470A	PPB	17 Dec 2021 12:21:55	S	Std
7.5 - 1	57.5	30335	-	7.5000	7470A	PPB	17 Dec 2021 12:24:11	S	Std
10.0 - 1	510	39783	-	10.0000	7470A	PPB	17 Dec 2021 12:29:01	S	Std
ICV16 - 1	ICV16	15736	3.8277	-	7470A	PPB	17 Dec 2021 12:31:41	U	SMPL
ICB16 - 1	ICB16	489	-0.0461	-	7470A	PPB	17 Dec 2021 12:33:56	U	SMPL
CCV43 - 1	CCV43	20616	5.0676	-	7470A	PPB	17 Dec 2021 12:36:11	U	SMPL
CCB43 - 1	CCB43	477	-0.0492	-	7470A	PPB	17 Dec 2021 12:38:26	U	SMPL
CRA - 1	CRA	1509	0.2130	-	7470A	PPB	17 Dec 2021 12:40:42	U	SMPL
HighStd - 1	HighStd	39659	9.9060	-	7470A	PPB	17 Dec 2021 12:42:58	U	SMPL
ChkStd - 1	ChkStd	27235	6.7493	-	7470A	PPB	17 Dec 2021 12:45:13	U	SMPL
PB141492BL - 1	PBW	949	0.0707	-	7470A	PPB	17 Dec 2021 12:47:29	U	SMPL
PB141492BS - 1	LCSW	13748	3.3226	-	7470A	PPB	17 Dec 2021 12:49:44	U	SMPL
M5051-05 - 1	P-1	289557	73.3987	-	7470A	PPB	17 Dec 2021 12:54:34	U	SMPL
M5051-06 - 1	P-2	804	0.0339	-	7470A	PPB	17 Dec 2021 12:56:50	U	SMPL
M5051-07 - 1	P-3	1549	0.2232	-	7470A	PPB	17 Dec 2021 12:59:10	U	SMPL
M5051-08 - 1	A5	797	0.0321	-	7470A	PPB	17 Dec 2021 13:01:27	U	SMPL
M5097-02 - 1	ETGI-371	470	-0.0510	-	7470A	PPB	17 Dec 2021 13:06:33	U	SMPL
CCV44 - 1	CCV44	20703	5.0897	-	7470A	PPB	17 Dec 2021 13:08:47	U	SMPL
CCB44 - 1	CCB44	464	-0.0525	-	7470A	PPB	17 Dec 2021 13:11:03	U	SMPL
M5097-02DUP - 1	ETGI-371DUP	719	0.0123	-	7470A	PPB	17 Dec 2021 13:13:18	U	SMPL
M5097-02MS - 1	ETGI-371MS	13816	3.3399	-	7470A	PPB	17 Dec 2021 13:15:34	U	SMPL
M5097-02MSD - 1	ETGI-371MSD	13704	3.3115	-	7470A	PPB	17 Dec 2021 13:17:49	U	SMPL
PB141492TB - 1	PB141454TB	921	0.0636	-	7470A	PPB	17 Dec 2021 13:20:05	U	SMPL
M5097-02LX5 - 1		977	0.0779	-	7470A	PPB	17 Dec 2021 13:22:21	U	SMPL
M5097-02A - 1		16036	3.9040	-	7470A	PPB	17 Dec 2021 13:24:37	U	SMPL
M5051-05DLX10 - 1		34064	8.4844	-	7470A	PPB	17 Dec 2021 13:26:54	U	SMPL
CCV45 - 1	CCV45	20490	5.0356	-	7470A	PPB	17 Dec 2021 13:29:11	U	SMPL
CCB45 - 1	CCB45	690	0.0049	-	7470A	PPB	17 Dec 2021 13:31:28	U	SMPL