



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117849

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/17/2021	08:46
2	CAL2	1	Water	NA	NA	20.3	7.00	12/17/2021	08:47
3	CAL3	1	Water	NA	NA	20.3	10.03	12/17/2021	08:49
4	ICV	1	Water	NA	NA	20.4	7.00	12/17/2021	08:50
5	CCV1	1	Water	NA	NA	20.4	2.01	12/17/2021	08:51
6	M5128-01	1	Water	NA	NA	20.2	6.01	12/17/2021	08:53
7	M5128-01DUP	1	Water	NA	NA	20.3	6.02	12/17/2021	08:54
8	M5128-03	1	Water	NA	NA	20.3	6.82	12/17/2021	08:56
9	M5128-05	1	Water	NA	NA	20.2	6.87	12/17/2021	09:00
10	M5128-07	1	Water	NA	NA	20.3	6.83	12/17/2021	09:01
11	M5135-01	1	Water	NA	NA	20.4	7.14	12/17/2021	09:03
12	CCV2	1	Water	NA	NA	20.6	12.02	12/17/2021	09:05

WORKLIST(Hardcopy Internal Chain)

117849

WorkList Name : PH M5135 W

WorkList ID : 155267

Department : Wet-Chemistry

Date : 12-17-2021 08:40:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/15/2021	Water	M5128-01	pH	Cool 4 deg C	LOCK01	H41	MW-1	12/15/2021	9040C
12/15/2021	Water	M5128-03	pH	Cool 4 deg C	LOCK01	H41	MW-2	12/15/2021	9040C
12/15/2021	Water	M5128-05	pH	Cool 4 deg C	LOCK01	H41	MW-3	12/15/2021	9040C
12/15/2021	Water	M5128-07	pH	Cool 4 deg C	LOCK01	H41	MW-4	12/15/2021	9040C
12/16/2021	Water	M5135-01	pH	Cool 4 deg C	PSEG03	H41	34009	12/16/2021	9040C

Date/Time 12-17-21 08:45

Raw Sample Received by: 201620

Raw Sample Relinquished by: 201620

Date/Time 12-17-21

Raw Sample Received by: 201620

Raw Sample Relinquished by: 201620