



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117853

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/17/2021	13:15
2	CAL2	1	Water	NA	NA	20.3	7.00	12/17/2021	13:16
3	CAL3	1	Water	NA	NA	20.3	10.03	12/17/2021	13:19
4	ICV	1	Water	NA	NA	20.4	7.00	12/17/2021	13:20
5	CCV1	1	Water	NA	NA	20.4	2.01	12/17/2021	13:22
6	M5145-01	1	Water	NA	NA	20.6	5.26	12/17/2021	13:25
7	M5145-01DUP	1	Water	NA	NA	20.7	5.27	12/17/2021	13:26
8	CCV2	1	Water	NA	NA	20.4	12.03	12/17/2021	13:29

13117853

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M5145 W WorkList ID : 155303 Department : Wet-Chemistry Date : 12-17-2021 13:01:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/15/2021	Water	M5145-01	pH	Cool 4 deg C	STER12	H11	CELL-POOL-WATER-4TH	12/15/2021	9040C

Date/Time 12.17.2021 13:10
Raw Sample Received by: JP (WCC)
Raw Sample Relinquished by: CP sm

Date/Time 12.17.2021 14:20
Raw Sample Received by: CP sm
Raw Sample Relinquished by: JP (WCC)