

LB 117856


SFAM01.1

INSTRUMENT 110: CV1

Linear

 μ Abs.:

44430



A= 0.0000e+000

B= 2.2520e-004 slope

C= -2.7267e-002 y-intercept

Rho= 0.9999647

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	401)
0.00	0.000	0.002	0.002	131	0.000	131					=
0.05	0.050					205	0				-16
0.20	0.200	0.168	-0.032	869	0.0 %	869					0
2.50	2.500	2.501	0.001	11227	0.0 %	11227					1
5.00	5.000	5.064	0.064	22608	0.0 %	22608					0
7.50	7.500	7.486	-0.014	33364	0.0 %	33364					0
10.0	10.000	9.978	-0.022	44430	0.0 %	44430					0

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	131	-		0 SFAM01.1	PPB	12/17/2021 14:29	Std
	0.2 S01	869	-		0.2 SFAM01.1	PPB	12/17/2021 14:32	Std
	2.5 S02	11227	-		2.5 SFAM01.1	PPB	12/17/2021 14:34	Std
	5 S03	22608	-		5 SFAM01.1	PPB	12/17/2021 14:36	Std
	7.5 S04	33364	-		7.5 SFAM01.1	PPB	12/17/2021 14:38	Std
	10 S05	44430	-		10 SFAM01.1	PPB	12/17/2021 14:41	Std
ICV098	ICV098	17757	3.9715 -		SFAM01.1	PPB	12/17/2021 14:44	SMPL
ICB098	ICB098	80	-0.0093 -		SFAM01.1	PPB	12/17/2021 14:46	SMPL
CCV046	CCV046	22207	4.9737 -		SFAM01.1	PPB	12/17/2021 14:48	SMPL
CCB046	CCB046	91	-0.0068 -		SFAM01.1	PPB	12/17/2021 14:50	SMPL
PB141519BL	PBS519	25	-0.0216 -		SFAM01.1	PPB	12/17/2021 14:55	SMPL
M5000-01	A0BA4	3401	0.7386 -		SFAM01.1	PPB	12/17/2021 14:57	SMPL
M5000-02	A0BA4D	4066	0.8884 -		SFAM01.1	PPB	12/17/2021 15:00	SMPL
M5000-03	A0BA4S	13322	2.9728 -		SFAM01.1	PPB	12/17/2021 15:02	SMPL
M5000-04	A0BA5	3314	0.719 -		SFAM01.1	PPB	12/17/2021 15:04	SMPL
M5000-05	A0BA6	414	0.066 -		SFAM01.1	PPB	12/17/2021 15:06	SMPL
M5000-06	A0BA7	413	0.0657 -		SFAM01.1	PPB	12/17/2021 15:09	SMPL
M5000-07	A0BA8	1255	0.2554 -		SFAM01.1	PPB	12/17/2021 15:11	SMPL
M5000-08	A0BA9	486	0.0822 -		SFAM01.1	PPB	12/17/2021 15:13	SMPL
M5000-09	A0BB0	7320	1.6212 -		SFAM01.1	PPB	12/17/2021 15:16	SMPL
M5000-10	A0BB1	374	0.057 -		SFAM01.1	PPB	12/17/2021 15:18	SMPL
M5000-11	A0BB2	522	0.0903 -		SFAM01.1	PPB	12/17/2021 15:20	SMPL
M5000-12	A0BB3	403	0.0635 -		SFAM01.1	PPB	12/17/2021 15:22	SMPL
M5000-13	A0BC4	18215	4.0747 -		SFAM01.1	PPB	12/17/2021 15:25	SMPL
M5000-14	A0BC5	519	0.0896 -		SFAM01.1	PPB	12/17/2021 15:27	SMPL
M5000-15	A0BC6	1190	0.2407 -		SFAM01.1	PPB	12/17/2021 15:29	SMPL
M5000-16	A0BC7	143656	32.3235 -		SFAM01.1	PPB	12/17/2021 15:31	SMPL
M5000-17	A0BC8	16361	3.6572 -		SFAM01.1	PPB	12/17/2021 15:34	SMPL
M5000-19	A0BD1	6779	1.4993 -		SFAM01.1	PPB	12/17/2021 15:36	SMPL
PB141520BL	PBS520	381	0.0585 -		SFAM01.1	PPB	12/17/2021 15:38	SMPL
CCV047	CCV047	21945	4.9147 -		SFAM01.1	PPB	12/17/2021 15:41	SMPL
CCB047	CCB047	66	-0.0124 -		SFAM01.1	PPB	12/17/2021 15:43	SMPL
M5001-01	A0BB4	3934	0.8587 -		SFAM01.1	PPB	12/17/2021 15:45	SMPL
M5001-02	A0BB5	447	0.0734 -		SFAM01.1	PPB	12/17/2021 15:47	SMPL
M5001-03	A0BB6	3377	0.7332 -		SFAM01.1	PPB	12/17/2021 15:50	SMPL
M5001-04	A0BB7	1205	0.2441 -		SFAM01.1	PPB	12/17/2021 15:52	SMPL
M5001-05	A0BB8	159269	35.8395 -		SFAM01.1	PPB	12/17/2021 15:54	SMPL
M5001-06	A0BB9	295	0.0392 -		SFAM01.1	PPB	12/17/2021 15:56	SMPL
M5001-07	A0BC0	322174	72.525 -		SFAM01.1	PPB	12/17/2021 15:59	SMPL
M5001-08	A0BC0D	306300	HIGH -		SFAM01.1	PPB	12/17/2021 16:01	
M5001-09	A0BC0S	339332	76.3889 -		SFAM01.1	PPB	12/17/2021 16:03	SMPL
M5001-10	A0BC1	7	-0.0257 -		SFAM01.1	PPB	12/17/2021 16:05	SMPL
M5001-11	A0BC2	4116	0.8996 -		SFAM01.1	PPB	12/17/2021 16:07	SMPL
M5001-12	A0BC3	444	0.0727 -		SFAM01.1	PPB	12/17/2021 16:10	SMPL
M5001-14	A0BD2	31919	7.1608 -		SFAM01.1	PPB	12/17/2021 16:12	SMPL
M5001-15	A0BD9	3774	0.8226 -		SFAM01.1	PPB	12/17/2021 16:14	SMPL

LB117856

INSTRUMENT ID : CV1

M5001-16	A0BE0	428	0.0691 -	SFAM01.1	PPB	12/17/2021 16:16 SMPL
PB141521BL	PBS521	-14	-0.0304 -	SFAM01.1	PPB	12/17/2021 16:19 SMPL
M5020-01	BGBL8	1824	0.3835 -	SFAM01.1	PPB	12/17/2021 16:21 SMPL
M5020-02	BGBL8D	1847	0.3887 -	SFAM01.1	PPB	12/17/2021 16:23 SMPL
M5020-03	BGBL8S	10687	2.3794 -	SFAM01.1	PPB	12/17/2021 16:26 SMPL
M5020-04	BGBL9	1645	0.3432 -	SFAM01.1	PPB	12/17/2021 16:28 SMPL
CCV048	CCV048	22993	5.1507 -	SFAM01.1	PPB	12/17/2021 16:33 SMPL
CCB048	CCB048	12	-0.0246 -	SFAM01.1	PPB	12/17/2021 16:35 SMPL
M5020-05	BGBM0	1341	0.2747 -	SFAM01.1	PPB	12/17/2021 16:37 SMPL
M5020-06	BGBM1	1086	0.2173 -	SFAM01.1	PPB	12/17/2021 16:40 SMPL
M5020-07	BGBM2	1325	0.2711 -	SFAM01.1	PPB	12/17/2021 16:42 SMPL
PB141522BL	PBS522	9	-0.0252 -	SFAM01.1	PPB	12/17/2021 16:44 SMPL
M5021-01	BGBM3	349	0.0513 -	SFAM01.1	PPB	12/17/2021 16:47 SMPL
M5021-02	BGBM4	360	0.0538 -	SFAM01.1	PPB	12/17/2021 16:49 SMPL
M5021-03	BGBM5	316	0.0439 -	SFAM01.1	PPB	12/17/2021 16:51 SMPL
M5021-04	BGBM6	410	0.0651 -	SFAM01.1	PPB	12/17/2021 16:53 SMPL
M5021-05	BGBM7	465	0.0774 -	SFAM01.1	PPB	12/17/2021 16:56 SMPL
M5021-06	BGBM8	509	0.0874 -	SFAM01.1	PPB	12/17/2021 16:58 SMPL
M5021-07	BGBM9	303	0.041 -	SFAM01.1	PPB	12/17/2021 17:00 SMPL
M5021-08	BGBN0	305	0.0414 -	SFAM01.1	PPB	12/17/2021 17:02 SMPL
M5021-09	BGBN0D	322	0.0452 -	SFAM01.1	PPB	12/17/2021 17:05 SMPL
M5021-10	BGBN0S	8045	1.7844 -	SFAM01.1	PPB	12/17/2021 17:07 SMPL
M5021-11	BGBN1	327	0.0464 -	SFAM01.1	PPB	12/17/2021 17:09 SMPL
M5021-12	MB0AB5	358	0.0534 -	SFAM01.1	PPB	12/17/2021 17:11 SMPL
CCV049	CCV049	22868	5.1225 -	SFAM01.1	PPB	12/17/2021 17:14 SMPL
CCB049	CCB049	50	-0.016 -	SFAM01.1	PPB	12/17/2021 17:16 SMPL