



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB117865

Slope : 98.7

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	12/17/2021	17:01
2	CAL2	1	Water	NA	NA	20.6	7.00	12/17/2021	17:02
3	CAL3	1	Water	NA	NA	20.6	10.03	12/17/2021	17:03
4	ICV	1	Water	NA	NA	20.5	7.02	12/17/2021	17:05
5	CCV1	1	Water	NA	NA	20.5	2.01	12/17/2021	17:06
6	M5142-02	1	Solid	20.03	20	21.2	8.89	12/17/2021	17:10
7	M5142-02DUP	1	Solid	20.04	20	21.3	8.90	12/17/2021	17:11
8	CCV2	1	Water	NA	NA	20.9	12.03	12/17/2021	17:12

187117865

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M5142 WorkList ID : 155321 Department : Wet-Chemistry Date : 12-17-2021 16:52:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/17/2021	Solid	M5142-02	Corrosivity	Cool 4 deg C	ENTA05	H42	A2	12/17/2021	9045D

Date/Time 12.17.21 18:00
Raw Sample Received by: BTCP/21/lab2
Raw Sample Relinquished by: BTCP/21/lab2

Date/Time 12.17.21 18:00
Raw Sample Received by: BTCP/21/lab2
Raw Sample Relinquished by: BTCP/21/lab2