


LB117869

SFAM01.1 INSTRUMENT ID: CV4

 μ Abs.:

48088



Linear

A= 0.0000e+000

B= 2.0830e-004 slope

C= -4.2020e-002 y-intercept

Rho= 0.9999699

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	(61)
0.00	0.000	0.006	0.006	232	0.000	232					
0.05	0.050					205	0				-17
0.20	0.200	0.167	-0.033	1003	0.0 %	1003					0
2.50	2.500	2.499	-0.001	12200	0.0 %	12200					0
5.00	5.000	5.056	0.056	24474	0.0 %	24474					1
7.50	7.500	7.497	-0.003	36192	0.0 %	36192					0
10.0	10.000	9.975	-0.025	48088	0.0 %	48088					0

LB117869

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	232	-		0 SFAM01.1	PPB	12/18/2021 8:52	Std
	0.2 S01	1003	-		0.2 SFAM01.1	PPB	12/18/2021 8:55	Std
	2.5 S02	12200	-		2.5 SFAM01.1	PPB	12/18/2021 8:57	Std
	5 S03	24474	-		5 SFAM01.1	PPB	12/18/2021 8:59	Std
	7.5 S04	36192	-		7.5 SFAM01.1	PPB	12/18/2021 9:01	Std
	10 S05	48088	-		10 SFAM01.1	PPB	12/18/2021 9:04	Std
ICV099	ICV099	18015	3.7105 -		SFAM01.1	PPB	12/18/2021 9:06	SMPL
ICB099	ICB099	65	-0.0285 -		SFAM01.1	PPB	12/18/2021 9:09	SMPL
CCV050	CCV050	23134	4.7768 -		SFAM01.1	PPB	12/18/2021 9:11	SMPL
CCB050	CCB050	27	-0.0364 -		SFAM01.1	PPB	12/18/2021 9:13	SMPL
M5000-16DLX5	A0BC7	30130	6.2341 -		SFAM01.1	PPB	12/18/2021 9:18	SMPL
M5001-05DLX10	A0BB8	17247	3.5506 -		SFAM01.1	PPB	12/18/2021 9:23	SMPL
M5001-07DLX25	A0BC0	14819	3.0448 -		SFAM01.1	PPB	12/18/2021 9:28	SMPL
M5001-08DLX25	A0BC0	15543	3.1956 -		SFAM01.1	PPB	12/18/2021 9:30	SMPL
M5001-09DLX25	A0BC0	15179	3.1198 -		SFAM01.1	PPB	12/18/2021 9:32	SMPL
CCV051	CCV051	25121	5.1907 -		SFAM01.1	PPB	12/18/2021 9:44	SMPL
CCB051	CCB051	63	-0.0289 -		SFAM01.1	PPB	12/18/2021 9:46	SMPL