

LB117871


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

39221



A= 0.0000e+000

B= 2.5618e-004 slope

C= 1.4358e-002 y-intercept

Rho= 0.9998990

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.060	0.060	178	0.000	178					=
0.05	0.050					205	0				-3
0.20	0.200	0.195	-0.005	706	0.0 %	706					-2
2.50	2.500	2.444	-0.056	9485	0.0 %	9485					0
5.00	5.000	5.014	0.014	19517	0.0 %	19517					-1
7.50	7.500	7.424	-0.076	28925	0.0 %	28925					-1
10.0	10.000	10.062	0.062	39221	0.0 %	39221					

LB117871

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	178	-		0 SFAM01.1	PPB	12/18/2021 14:14	Std
	0.2 S01	706	-		0.2 SFAM01.1	PPB	12/18/2021 14:16	Std
	2.5 S2	9485	-		2.5 SFAM01.1	PPB	12/18/2021 14:18	Std
	5 S03	19517	-		5 SFAM01.1	PPB	12/18/2021 14:20	Std
	7.5 S04	28925	-		7.5 SFAM01.1	PPB	12/18/2021 14:23	Std
	10 S05	39221	-		10 SFAM01.1	PPB	12/18/2021 14:25	Std
ICV001	ICV001	14929	3.8389 -		SFAM01.1	PPB	12/18/2021 14:28	SMPL
ICB001	ICB001	109	0.0423 -		SFAM01.1	PPB	12/18/2021 14:30	SMPL
CCV055	CCV055	19431	4.9922 -		SFAM01.1	PPB	12/18/2021 14:32	SMPL
CCB055	CCB055	106	0.0415 -		SFAM01.1	PPB	12/18/2021 14:34	SMPL
PB141555BL	PBW555	-18	0.0097 -		SFAM01.1	PPB	12/18/2021 14:37	SMPL
M5047-01	MH1E32	11	0.0172 -		SFAM01.1	PPB	12/18/2021 14:39	SMPL
M5047-02	MH1E36	158	0.0548 -		SFAM01.1	PPB	12/18/2021 14:41	SMPL
M5047-03	MH1E38	202	0.0661 -		SFAM01.1	PPB	12/18/2021 14:43	SMPL
M5047-04	MH1E40	167	0.0571 -		SFAM01.1	PPB	12/18/2021 14:46	SMPL
M5047-05	MH1E42	280	0.0861 -		SFAM01.1	PPB	12/18/2021 14:48	SMPL
M5047-06	MH1E44	514	0.146 -		SFAM01.1	PPB	12/18/2021 14:50	SMPL
M5047-07	MH1E46	119	0.0448 -		SFAM01.1	PPB	12/18/2021 14:53	SMPL
M5047-08	MH1E48	295	0.0899 -		SFAM01.1	PPB	12/18/2021 14:55	SMPL
M5047-09	MH1E50	303	0.092 -		SFAM01.1	PPB	12/18/2021 14:57	SMPL
M5047-10	MH1E52	5	0.0156 -		SFAM01.1	PPB	12/18/2021 14:59	SMPL
M5047-11	MH1E54	175	0.0592 -		SFAM01.1	PPB	12/18/2021 15:02	SMPL
M5047-12	MH1E56	151	0.053 -		SFAM01.1	PPB	12/18/2021 15:04	SMPL
M5047-13	MH1E58	320	0.0963 -		SFAM01.1	PPB	12/18/2021 15:06	SMPL
M5047-14	MH1E60	417	0.1212 -		SFAM01.1	PPB	12/18/2021 15:08	SMPL
M5047-15	MH1E62	23	0.0202 -		SFAM01.1	PPB	12/18/2021 15:11	SMPL
M5047-16	MH1E64	245	0.0771 -		SFAM01.1	PPB	12/18/2021 15:13	SMPL
M5047-17	MH1E66	46	0.0261 -		SFAM01.1	PPB	12/18/2021 15:15	SMPL
M5047-18	MH1E68	81	0.0351 -		SFAM01.1	PPB	12/18/2021 15:17	SMPL
M5047-19	MH1E70	91	0.0377 -		SFAM01.1	PPB	12/18/2021 15:20	SMPL
CCV056	CCV056	19595	5.0342 -		SFAM01.1	PPB	12/18/2021 15:22	SMPL
CCB056	CCB056	106	0.0415 -		SFAM01.1	PPB	12/18/2021 15:24	SMPL
M5047-20	MH1E72	483	0.1381 -		SFAM01.1	PPB	12/18/2021 15:27	SMPL
M5047-21	MH1E72D	468	0.1343 -		SFAM01.1	PPB	12/18/2021 15:29	SMPL
M5047-22	MH1E72S	3959	1.0286 -		SFAM01.1	PPB	12/18/2021 15:31	SMPL
PB141556BL	PBW556	10	0.0169 -		SFAM01.1	PPB	12/18/2021 15:36	SMPL
M5048-01	MH1E74	413	0.1202 -		SFAM01.1	PPB	12/18/2021 15:38	SMPL
M5048-02	MH1E76	362	0.1071 -		SFAM01.1	PPB	12/18/2021 15:40	SMPL
M5048-03	MH1E78	345	0.1027 -		SFAM01.1	PPB	12/18/2021 15:43	SMPL
M5048-04	MH1E80	412	0.1199 -		SFAM01.1	PPB	12/18/2021 15:45	SMPL
M5048-05	MH1E82	291	0.0889 -		SFAM01.1	PPB	12/18/2021 15:47	SMPL
M5048-06	MH1E84	359	0.1063 -		SFAM01.1	PPB	12/18/2021 15:50	SMPL
M5048-07	MH1E86	365	0.1079 -		SFAM01.1	PPB	12/18/2021 15:52	SMPL
M5048-08	MH1E88	370	0.1091 -		SFAM01.1	PPB	12/18/2021 15:54	SMPL
M5048-09	MH1E90	644	0.1793 -		SFAM01.1	PPB	12/18/2021 15:56	SMPL
M5048-10	MH1E92	387	0.1135 -		SFAM01.1	PPB	12/18/2021 15:59	SMPL

LB117871

INSTRUMENT ID : CV1

M5048-11	MH1E94	380	0.1117 -	SFAM01.1	PPB	12/18/2021 16:01 SMPL
M5048-12	MH1E96	418	0.1214 -	SFAM01.1	PPB	12/18/2021 16:03 SMPL
M5048-13	MH1E98	527	0.1494 -	SFAM01.1	PPB	12/18/2021 16:05 SMPL
M5048-14	MH1F00	501	0.1427 -	SFAM01.1	PPB	12/18/2021 16:08 SMPL
M5048-15	MH1F02	509	0.1448 -	SFAM01.1	PPB	12/18/2021 16:10 SMPL
M5048-16	MH1F04	678	0.188 -	SFAM01.1	PPB	12/18/2021 16:12 SMPL
CCV057	CCV057	19605	5.0368 -	SFAM01.1	PPB	12/18/2021 16:14 SMPL
CCB057	CCB057	127	0.0469 -	SFAM01.1	PPB	12/18/2021 16:17 SMPL
M5048-17	MH1F06	617	0.1724 -	SFAM01.1	PPB	12/18/2021 16:19 SMPL
M5048-18	MH1F08	597	0.1673 -	SFAM01.1	PPB	12/18/2021 16:21 SMPL
M5048-19	MH1F10	825	0.2257 -	SFAM01.1	PPB	12/18/2021 16:24 SMPL
M5048-20	MH1F12	1162	0.312 -	SFAM01.1	PPB	12/18/2021 16:26 SMPL
M5048-21	MH1F12D	1187	0.3184 -	SFAM01.1	PPB	12/18/2021 16:28 SMPL
M5048-22	MH1F12S	4650	1.2056 -	SFAM01.1	PPB	12/18/2021 16:30 SMPL
PB141557BL	PBW557	-16	0.0103 -	SFAM01.1	PPB	12/18/2021 16:33 SMPL
M5049-01	MH1F14	768	0.2111 -	SFAM01.1	PPB	12/18/2021 16:35 SMPL
M5049-02	MH1F16	356	0.1056 -	SFAM01.1	PPB	12/18/2021 16:37 SMPL
M5049-03	MH1F18	1573	0.4173 -	SFAM01.1	PPB	12/18/2021 16:39 SMPL
M5049-04	MH1F20	1656	0.4386 -	SFAM01.1	PPB	12/18/2021 16:42 SMPL
M5049-05	MH1F22	999	0.2703 -	SFAM01.1	PPB	12/18/2021 16:44 SMPL
M5049-06	MH1F24	1856	0.4898 -	SFAM01.1	PPB	12/18/2021 16:46 SMPL
M5049-07	MH1F26	1979	0.5213 -	SFAM01.1	PPB	12/18/2021 16:49 SMPL
M5049-08	MH1F28	1517	0.403 -	SFAM01.1	PPB	12/18/2021 16:51 SMPL
M5049-09	MH1F30	3077	0.8026 -	SFAM01.1	PPB	12/18/2021 16:53 SMPL
M5049-10	MH1F32	207	0.0674 -	SFAM01.1	PPB	12/18/2021 16:55 SMPL
M5049-11	MH1F34	215	0.0694 -	SFAM01.1	PPB	12/18/2021 16:58 SMPL
M5049-12	MH1F36	1057	0.2851 -	SFAM01.1	PPB	12/18/2021 17:00 SMPL
M5049-13	MH1F38	1125	0.3026 -	SFAM01.1	PPB	12/18/2021 17:02 SMPL
CCV058	CCV058	19641	5.046 -	SFAM01.1	PPB	12/18/2021 17:04 SMPL
CCB058	CCB058	97	0.0392 -	SFAM01.1	PPB	12/18/2021 17:07 SMPL
M5049-14	MH1F40	789	0.2165 -	SFAM01.1	PPB	12/18/2021 17:09 SMPL
M5049-15	MH1F42	196	0.0646 -	SFAM01.1	PPB	12/18/2021 17:11 SMPL
M5049-16	MH1F44	716	0.1978 -	SFAM01.1	PPB	12/18/2021 17:13 SMPL
M5049-17	MH1F46	194	0.0641 -	SFAM01.1	PPB	12/18/2021 17:16 SMPL
M5049-18	MH1F48	215	0.0694 -	SFAM01.1	PPB	12/18/2021 17:18 SMPL
M5049-19	MH1F50	228	0.0728 -	SFAM01.1	PPB	12/18/2021 17:20 SMPL
M5049-20	MH1F52	177	0.0597 -	SFAM01.1	PPB	12/18/2021 17:23 SMPL
M5049-21	MH1F52D	178	0.06 -	SFAM01.1	PPB	12/18/2021 17:25 SMPL
M5049-22	MH1F52S	3681	0.9574 -	SFAM01.1	PPB	12/18/2021 17:27 SMPL
PB141558BL	PBW558	-9	0.0121 -	SFAM01.1	PPB	12/18/2021 17:29 SMPL
M5050-01	MH1F54	380	0.1117 -	SFAM01.1	PPB	12/18/2021 17:32 SMPL
M5050-02	MH1F56	258	0.0805 -	SFAM01.1	PPB	12/18/2021 17:34 SMPL
M5050-03	MH1F58	115	0.0438 -	SFAM01.1	PPB	12/18/2021 17:36 SMPL
M5050-04	MH1F60	-1	0.0141 -	SFAM01.1	PPB	12/18/2021 17:38 SMPL
M5050-05	MH1F62	-24	0.0082 -	SFAM01.1	PPB	12/18/2021 17:41 SMPL
M5050-06	MH1F64	-18	0.0097 -	SFAM01.1	PPB	12/18/2021 17:43 SMPL
M5050-07	MH1F66	2	0.0149 -	SFAM01.1	PPB	12/18/2021 17:45 SMPL

LB117871
INSTRUMENT ID : CV1

M5050-08	MH1F68	-5	0.0131	-	SFAM01.1	PPB	12/18/2021 17:48	SMPL
M5050-09	MH1F70	-10	0.0118	-	SFAM01.1	PPB	12/18/2021 17:50	SMPL
M5050-10	MH1F72	-14	0.0108	-	SFAM01.1	PPB	12/18/2021 17:52	SMPL
CCV059	CCV059	19770	5.079	-	SFAM01.1	PPB	12/18/2021 17:54	SMPL
CCB059	CCB059	140	0.0502	-	SFAM01.1	PPB	12/18/2021 17:57	SMPL
M5050-11	MH1F74	15	0.0182	-	SFAM01.1	PPB	12/18/2021 17:59	SMPL
M5050-12	MH1F76	280	0.0861	-	SFAM01.1	PPB	12/18/2021 18:01	SMPL
M5050-13	MH1F78	259	0.0807	-	SFAM01.1	PPB	12/18/2021 18:04	SMPL
M5050-14	MH1F80	348	0.1035	-	SFAM01.1	PPB	12/18/2021 18:06	SMPL
M5050-15	MH1F82	463	0.133	-	SFAM01.1	PPB	12/18/2021 18:08	SMPL
M5050-16	MH1F84	451	0.1299	-	SFAM01.1	PPB	12/18/2021 18:10	SMPL
M5050-17	MH1F86	892	0.2429	-	SFAM01.1	PPB	12/18/2021 18:13	SMPL
M5050-18	MH1F88	1990	0.5242	-	SFAM01.1	PPB	12/18/2021 18:15	SMPL
M5050-19	MH1F88D	1959	0.5162	-	SFAM01.1	PPB	12/18/2021 18:17	SMPL
M5050-20	MH1F88S	5853	1.5138	-	SFAM01.1	PPB	12/18/2021 18:19	SMPL
CCV060	CCV060	19734	5.0698	-	SFAM01.1	PPB	12/18/2021 18:22	SMPL
CCB060	CCB060	125	0.0464	-	SFAM01.1	PPB	12/18/2021 18:24	SMPL