



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 12/20/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 15:00
In Date: 12/19/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 102
Time OUT: 08:00
Out Date: 12/20/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117872

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M5156-01	EX4Y3	1	1.14	8.40	9.54	7.78	79.0	
M5156-02	EX4Y4	2	1.14	8.74	9.88	8.08	79.4	
M5156-03	EX4Y5	3	1.15	8.83	9.98	8.51	83.4	
M5156-04	EX4Y6	4	1.16	8.58	9.74	8.16	81.6	
M5156-05	EX4Y7	5	1.14	8.80	9.94	8.46	83.2	
M5156-06	EX4Y8	6	1.12	8.38	9.5	8.00	82.1	
M5156-07	EX4Y9	7	1.18	8.46	9.64	8.11	81.9	
M5156-08	EX4Z0	8	1.15	8.78	9.93	8.51	83.8	
M5156-09	EX4Z1	9	1.17	8.35	9.52	7.92	80.8	
M5156-10	EX4Z2	10	1.18	8.46	9.64	7.86	79.0	
M5156-11	EX503	11	1.14	8.50	9.64	8.21	83.2	
M5156-12	EX504	12	1.15	8.81	9.96	8.31	81.3	
M5156-13	EX505	13	1.17	8.59	9.76	8.25	82.4	
M5156-14	EX505D	14	1.17	8.59	9.76	8.25	82.4	
M5156-15	EX505S	15	1.17	8.59	9.76	8.25	82.4	
M5156-16	EX506	16	1.19	8.63	9.82	8.53	85.1	
M5156-17	EX507	17	1.19	8.53	9.72	8.1	81.0	
M5156-18	EX508	18	1.19	8.63	9.82	8.22	81.5	
M5156-19	EX509	19	1.15	8.83	9.98	8.23	80.2	
M5156-20	EX510	20	1.18	8.76	9.94	8.44	82.9	
M5156-21	EX511	21	1.14	8.58	9.72	8.29	83.3	
M5156-22	EX512	22	1.19	8.60	9.79	8.12	80.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$