



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 12/20/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 15:19
In Date: 12/19/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 102
Time OUT: 08:05
Out Date: 12/20/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117873

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M5157-01	EX4Z3	1	1.15	8.40	9.55	8.25	84.5	
M5157-02	EX4Z4	2	1.18	8.51	9.69	8.23	82.8	
M5157-03	EX4Z5	3	1.18	8.35	9.53	8.18	83.8	
M5157-04	EX4Z6	4	1.19	8.37	9.56	7.95	80.8	
M5157-05	EX4Z7	5	1.16	8.80	9.96	8.71	85.8	
M5157-06	EX4Z8	6	1.18	8.80	9.98	8.77	86.3	
M5157-07	EX4Z9	7	1.15	8.67	9.82	8.35	83.0	
M5157-08	EX500	8	1.11	8.71	9.82	8.44	84.2	
M5157-09	EX501	9	1.13	8.39	9.52	8.2	84.3	
M5157-10	EX502	10	1.19	8.49	9.68	8.17	82.2	
M5157-11	EX513	11	1.18	8.71	9.89	8.2	80.6	
M5157-12	EX514	12	1.15	8.81	9.96	8.39	82.2	
M5157-13	EX515	13	1.14	8.74	9.88	7.83	76.5	
M5157-14	EX516	14	1.11	8.71	9.82	8.11	80.4	
M5157-15	EX517	15	1.11	8.53	9.64	7.74	77.7	
M5157-16	EX518	16	1.11	8.71	9.82	8.33	82.9	
M5157-17	EX519	17	1.19	8.67	9.86	7.67	74.7	
M5157-18	EX520	18	1.18	8.37	9.55	7.76	78.6	
M5157-19	EX521	19	1.14	8.83	9.97	8.48	83.1	
M5157-20	EX522	20	1.15	8.73	9.88	8.34	82.4	
M5157-21	EX522D	21	1.15	8.73	9.88	8.34	82.4	
M5157-22	EX522S	22	1.15	8.73	9.88	8.34	82.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$