

LB117876


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

39480



A= 0.0000e+000

B= 2.5209e-004 slope

C= 1.7419e-002 y-intercept

Rho= 0.9999081

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.053	0.053	142	0.000	142					
0.05	0.050					205	0				
0.20	0.200	0.193	-0.007	698	0.0 %	698					-4
2.50	2.500	2.427	-0.073	9560	0.0 %	9560					-3
5.00	5.000	4.981	-0.019	19689	0.0 %	19689					0
7.50	7.500	7.575	0.075	29979	0.0 %	29979					1
10.0	10.000	9.970	-0.030	39480	0.0 %	39480					0

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	142	-		0 SFAM01.1	PPB	12/19/2021 11:51	Std
	0.2 S01	698	-		0.2 SFAM01.1	PPB	12/19/2021 11:53	Std
	2.5 S02	9560	-		2.5 SFAM01.1	PPB	12/19/2021 11:55	Std
	5 S03	19689	-		5 SFAM01.1	PPB	12/19/2021 11:57	Std
	7.5 S04	29979	-		7.5 SFAM01.1	PPB	12/19/2021 12:00	Std
	10 S05	39480	-		10 SFAM01.1	PPB	12/19/2021 12:05	Std
ICV002	ICV002	15783	3.9962	-	SFAM01.1	PPB	12/19/2021 12:07	SMPL
ICB002	ICB002	109	0.0449	-	SFAM01.1	PPB	12/19/2021 12:10	SMPL
CCV061	CCV061	19688	4.9807	-	SFAM01.1	PPB	12/19/2021 12:12	SMPL
CCB061	CCB061	-33	0.0091	-	SFAM01.1	PPB	12/19/2021 12:14	SMPL
PB141563BL	PBW563	6	0.0189	-	SFAM01.1	PPB	12/19/2021 12:16	SMPL
M5052-01	MH1D93	-5	0.0162	-	SFAM01.1	PPB	12/19/2021 12:19	SMPL
M5052-02	MH1D95	4	0.0184	-	SFAM01.1	PPB	12/19/2021 12:21	SMPL
M5052-03	MH1D97	1	0.0177	-	SFAM01.1	PPB	12/19/2021 12:23	SMPL
M5052-04	MH1D99	29	0.0247	-	SFAM01.1	PPB	12/19/2021 12:25	SMPL
M5052-05	MH1E01	9	0.0197	-	SFAM01.1	PPB	12/19/2021 12:28	SMPL
M5052-06	MH1E03	4	0.0184	-	SFAM01.1	PPB	12/19/2021 12:30	SMPL
M5052-07	MH1E05	3	0.0182	-	SFAM01.1	PPB	12/19/2021 12:32	SMPL
M5052-08	MH1E07	9	0.0197	-	SFAM01.1	PPB	12/19/2021 12:35	SMPL
M5052-09	MH1E09	-8	0.0154	-	SFAM01.1	PPB	12/19/2021 12:37	SMPL
M5052-10	MH1E11	30	0.025	-	SFAM01.1	PPB	12/19/2021 12:39	SMPL
M5052-11	MH1E13	21	0.0227	-	SFAM01.1	PPB	12/19/2021 12:41	SMPL
M5052-12	MH1E15	27	0.0242	-	SFAM01.1	PPB	12/19/2021 12:44	SMPL
M5052-13	MH1E17	20	0.0225	-	SFAM01.1	PPB	12/19/2021 12:46	SMPL
M5052-14	MH1E19	24	0.0235	-	SFAM01.1	PPB	12/19/2021 12:48	SMPL
M5052-15	MH1E21	31	0.0252	-	SFAM01.1	PPB	12/19/2021 12:50	SMPL
M5052-16	MH1E23	14	0.0209	-	SFAM01.1	PPB	12/19/2021 12:53	SMPL
M5052-17	MH1E25	6	0.0189	-	SFAM01.1	PPB	12/19/2021 12:55	SMPL
M5052-18	MH1E27	31	0.0252	-	SFAM01.1	PPB	12/19/2021 12:57	SMPL
M5052-19	MH1E29	21	0.0227	-	SFAM01.1	PPB	12/19/2021 12:59	SMPL
CCV062	CCV062	19784	5.0049	-	SFAM01.1	PPB	12/19/2021 13:02	SMPL
CCB062	CCB062	98	0.0421	-	SFAM01.1	PPB	12/19/2021 13:04	SMPL
M5052-20	MH1E31	26	0.024	-	SFAM01.1	PPB	12/19/2021 13:06	SMPL
M5052-21	MH1E31D	22	0.023	-	SFAM01.1	PPB	12/19/2021 13:09	SMPL
M5052-22	MH1E31S	3607	0.9267	-	SFAM01.1	PPB	12/19/2021 13:11	SMPL
PB141564BL	PBW564	81	0.0378	-	SFAM01.1	PPB	12/19/2021 13:13	SMPL
M5053-01	MH1E33	18	0.022	-	SFAM01.1	PPB	12/19/2021 13:15	SMPL
M5053-02	MH1E37	38	0.027	-	SFAM01.1	PPB	12/19/2021 13:18	SMPL
M5053-03	MH1E39	21	0.0227	-	SFAM01.1	PPB	12/19/2021 13:20	SMPL
M5053-04	MH1E41	21	0.0227	-	SFAM01.1	PPB	12/19/2021 13:22	SMPL
M5053-05	MH1E43	28	0.0245	-	SFAM01.1	PPB	12/19/2021 13:24	SMPL
M5053-06	MH1E45	17	0.0217	-	SFAM01.1	PPB	12/19/2021 13:27	SMPL
M5053-07	MH1E47	35	0.0262	-	SFAM01.1	PPB	12/19/2021 13:29	SMPL
M5053-08	MH1E49	21	0.0227	-	SFAM01.1	PPB	12/19/2021 13:31	SMPL
M5053-09	MH1E51	22	0.023	-	SFAM01.1	PPB	12/19/2021 13:33	SMPL
M5053-10	MH1E53	11	0.0202	-	SFAM01.1	PPB	12/19/2021 13:36	SMPL

INSTRUMENT ID : CV1

M5053-11	MH1E55	11	0.0202 -	SFAM01.1	PPB	12/19/2021 13:38	SMPL
M5053-12	MH1E57	34	0.026 -	SFAM01.1	PPB	12/19/2021 13:40	SMPL
M5053-13	MH1E59	12	0.0204 -	SFAM01.1	PPB	12/19/2021 13:42	SMPL
M5053-14	MH1E61	8	0.0194 -	SFAM01.1	PPB	12/19/2021 13:45	SMPL
M5053-15	MH1E63	27	0.0242 -	SFAM01.1	PPB	12/19/2021 13:47	SMPL
M5053-16	MH1E65	66	0.0341 -	SFAM01.1	PPB	12/19/2021 13:49	SMPL
CCV063	CCV063	19910	5.0366 -	SFAM01.1	PPB	12/19/2021 13:52	SMPL
CCB063	CCB063	-43	0.0066 -	SFAM01.1	PPB	12/19/2021 13:54	SMPL
M5053-17	MH1E67	41	0.0278 -	SFAM01.1	PPB	12/19/2021 13:56	SMPL
M5053-18	MH1E69	34	0.026 -	SFAM01.1	PPB	12/19/2021 13:58	SMPL
M5053-19	MH1E71	19	0.0222 -	SFAM01.1	PPB	12/19/2021 14:01	SMPL
M5053-20	MH1E73	21	0.0227 -	SFAM01.1	PPB	12/19/2021 14:03	SMPL
M5053-21	MH1E73D	32	0.0255 -	SFAM01.1	PPB	12/19/2021 14:05	SMPL
M5053-22	MH1E73S	3397	0.8738 -	SFAM01.1	PPB	12/19/2021 14:07	SMPL
PB141565BL	PBW565	4	0.0184 -	SFAM01.1	PPB	12/19/2021 14:10	SMPL
M5054-01	MH1E75	19	0.0222 -	SFAM01.1	PPB	12/19/2021 14:12	SMPL
M5054-02	MH1E77	10	0.0199 -	SFAM01.1	PPB	12/19/2021 14:14	SMPL
M5054-03	MH1E79	12	0.0204 -	SFAM01.1	PPB	12/19/2021 14:16	SMPL
M5054-04	MH1E81	10	0.0199 -	SFAM01.1	PPB	12/19/2021 14:19	SMPL
M5054-05	MH1E83	21	0.0227 -	SFAM01.1	PPB	12/19/2021 14:21	SMPL
M5054-06	MH1E85	12	0.0204 -	SFAM01.1	PPB	12/19/2021 14:23	SMPL
M5054-07	MH1E87	13	0.0207 -	SFAM01.1	PPB	12/19/2021 14:26	SMPL
M5054-08	MH1E89	20	0.0225 -	SFAM01.1	PPB	12/19/2021 14:28	SMPL
M5054-09	MH1E91	47	0.0293 -	SFAM01.1	PPB	12/19/2021 14:30	SMPL
M5054-10	MH1E93	204	0.0688 -	SFAM01.1	PPB	12/19/2021 14:32	SMPL
M5054-11	MH1E95	174	0.0613 -	SFAM01.1	PPB	12/19/2021 14:35	SMPL
M5054-12	MH1E97	89	0.0399 -	SFAM01.1	PPB	12/19/2021 14:37	SMPL
M5054-13	MH1E99	92	0.0406 -	SFAM01.1	PPB	12/19/2021 14:39	SMPL
CCV064	CCV064	19890	5.0316 -	SFAM01.1	PPB	12/19/2021 14:41	SMPL
CCB064	CCB064	-54	0.0038 -	SFAM01.1	PPB	12/19/2021 14:44	SMPL
M5054-14	MH1F01	133	0.0509 -	SFAM01.1	PPB	12/19/2021 14:46	SMPL
M5054-15	MH1F03	92	0.0406 -	SFAM01.1	PPB	12/19/2021 14:48	SMPL
M5054-16	MH1F05	93	0.0409 -	SFAM01.1	PPB	12/19/2021 14:51	SMPL
M5054-17	MH1F07	111	0.0454 -	SFAM01.1	PPB	12/19/2021 14:53	SMPL
M5054-18	MH1F09	122	0.0482 -	SFAM01.1	PPB	12/19/2021 14:55	SMPL
M5054-19	MH1F11	149	0.055 -	SFAM01.1	PPB	12/19/2021 14:57	SMPL
M5054-20	MH1F13	103	0.0434 -	SFAM01.1	PPB	12/19/2021 15:00	SMPL
M5054-21	MH1F13D	112	0.0457 -	SFAM01.1	PPB	12/19/2021 15:02	SMPL
M5054-22	MH1F13S	4672	1.1952 -	SFAM01.1	PPB	12/19/2021 15:04	SMPL
PB141566BL	PBW566	13	0.0207 -	SFAM01.1	PPB	12/19/2021 15:06	SMPL
M5055-01	MH1F15	18	0.022 -	SFAM01.1	PPB	12/19/2021 15:09	SMPL
M5055-02	MH1F17	20	0.0225 -	SFAM01.1	PPB	12/19/2021 15:11	SMPL
M5055-03	MH1F19	22	0.023 -	SFAM01.1	PPB	12/19/2021 15:13	SMPL
M5055-04	MH1F21	16	0.0215 -	SFAM01.1	PPB	12/19/2021 15:16	SMPL
M5055-05	MH1F23	-3	0.0167 -	SFAM01.1	PPB	12/19/2021 15:18	SMPL
M5055-06	MH1F25	20	0.0225 -	SFAM01.1	PPB	12/19/2021 15:20	SMPL
M5055-07	MH1F27	34	0.026 -	SFAM01.1	PPB	12/19/2021 15:22	SMPL

INSTRUMENT ID : CV1

M5055-08	MH1F29	11	0.0202 -	SFAM01.1	PPB	12/19/2021 15:25	SMPL
M5055-09	MH1F31	18	0.022 -	SFAM01.1	PPB	12/19/2021 15:27	SMPL
M5055-10	MH1F33	9	0.0197 -	SFAM01.1	PPB	12/19/2021 15:29	SMPL
CCV065	CCV065	19687	4.9804 -	SFAM01.1	PPB	12/19/2021 15:31	SMPL
CCB065	CCB065	-47	0.0056 -	SFAM01.1	PPB	12/19/2021 15:34	SMPL
M5055-11	MH1F35	0	0.0174 -	SFAM01.1	PPB	12/19/2021 15:36	SMPL
M5055-12	MH1F37	39	0.0273 -	SFAM01.1	PPB	12/19/2021 15:38	SMPL
M5055-13	MH1F39	25	0.0237 -	SFAM01.1	PPB	12/19/2021 15:41	SMPL
M5055-14	MH1F41	29	0.0247 -	SFAM01.1	PPB	12/19/2021 15:43	SMPL
M5055-15	MH1F43	0	0.0174 -	SFAM01.1	PPB	12/19/2021 15:45	SMPL
M5055-16	MH1F45	22	0.023 -	SFAM01.1	PPB	12/19/2021 15:47	SMPL
M5055-17	MH1F47	11	0.0202 -	SFAM01.1	PPB	12/19/2021 15:50	SMPL
M5055-18	MH1F49	5	0.0187 -	SFAM01.1	PPB	12/19/2021 15:52	SMPL
M5055-19	MH1F51	48	0.0295 -	SFAM01.1	PPB	12/19/2021 15:54	SMPL
M5055-20	MH1F53	45	0.0288 -	SFAM01.1	PPB	12/19/2021 15:57	SMPL
M5055-21	MH1F53D	50	0.03 -	SFAM01.1	PPB	12/19/2021 15:59	SMPL
M5055-22	MH1F53S	4060	1.0409 -	SFAM01.1	PPB	12/19/2021 16:01	SMPL
PB141567BL	PBW567	11	0.0202 -	SFAM01.1	PPB	12/19/2021 16:06	SMPL
M5056-01	MH1F55	333	0.1014 -	SFAM01.1	PPB	12/19/2021 16:08	SMPL
M5056-02	MH1F57	265	0.0842 -	SFAM01.1	PPB	12/19/2021 16:11	SMPL
M5056-03	MH1F59	320	0.0981 -	SFAM01.1	PPB	12/19/2021 16:13	SMPL
M5056-04	MH1F61	311	0.0958 -	SFAM01.1	PPB	12/19/2021 16:15	SMPL
M5056-05	MH1F63	329	0.1004 -	SFAM01.1	PPB	12/19/2021 16:17	SMPL
M5056-06	MH1F65	288	0.09 -	SFAM01.1	PPB	12/19/2021 16:20	SMPL
M5056-07	MH1F67	287	0.0898 -	SFAM01.1	PPB	12/19/2021 16:22	SMPL
CCV066	CCV066	19650	4.9711 -	SFAM01.1	PPB	12/19/2021 16:28	SMPL
CCB066	CCB066	-36	0.0083 -	SFAM01.1	PPB	12/19/2021 16:31	SMPL
M5056-08	MH1F69	303	0.0938 -	SFAM01.1	PPB	12/19/2021 16:33	SMPL
M5056-09	MH1F71	280	0.088 -	SFAM01.1	PPB	12/19/2021 16:35	SMPL
M5056-10	MH1F73	321	0.0983 -	SFAM01.1	PPB	12/19/2021 16:37	SMPL
M5056-11	MH1F75	355	0.1069 -	SFAM01.1	PPB	12/19/2021 16:40	SMPL
M5056-12	MH1F77	450	0.1309 -	SFAM01.1	PPB	12/19/2021 16:42	SMPL
M5056-13	MH1F79	1441	0.3807 -	SFAM01.1	PPB	12/19/2021 16:44	SMPL
M5056-14	MH1F81	1506	0.3971 -	SFAM01.1	PPB	12/19/2021 16:46	SMPL
M5056-15	MH1F83	1575	0.4145 -	SFAM01.1	PPB	12/19/2021 16:49	SMPL
M5056-16	MH1F85	1530	0.4031 -	SFAM01.1	PPB	12/19/2021 16:51	SMPL
M5056-17	MH1F87	1468	0.3875 -	SFAM01.1	PPB	12/19/2021 16:53	SMPL
M5056-18	MH1F89	1722	0.4515 -	SFAM01.1	PPB	12/19/2021 16:56	SMPL
M5056-19	MH1F89D	1701	0.4462 -	SFAM01.1	PPB	12/19/2021 16:58	SMPL
M5056-20	MH1F89S	5221	1.3336 -	SFAM01.1	PPB	12/19/2021 17:00	SMPL
CCV067	CCV067	20741	5.2461 -	SFAM01.1	PPB	12/19/2021 17:02	SMPL
CCB067	CCB067	-60	0.0023 -	SFAM01.1	PPB	12/19/2021 17:05	SMPL