

LB 117882


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

51362



A= 0.0000e+000

B= 1.9526e-004

C= -1.0203e-002

Rho= 0.9999850

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.025	0.025	182	0.000	182					
0.2	0.200	0.208	0.008	1116	0.0 %	1116					4
2.5	2.500	2.474	-0.026	12723	0.0 %	12723					-1
5.0	5.000	4.973	-0.027	25521	0.0 %	25521					-1
7.5	7.500	7.501	0.001	38467	0.0 %	38467					0
10.0	10.000	10.019	0.019	51362	0.0 %	51362					0

LB117882 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 20 Dec 2021 09:34:37

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	182	-	0.00007471B	PPB		20 Dec 2021 09:35:34	S	Std
0.2 - 1	50.2	1116	-	0.20007471B	PPB		20 Dec 2021 09:37:52	S	Std
2.5 - 1	52.5	12723	-	2.50007471B	PPB		20 Dec 2021 09:40:08	S	Std
5.0 - 1	55	25521	-	5.00007471B	PPB		20 Dec 2021 09:42:25	S	Std
7.5 - 1	57.5	38467	-	7.50007471B	PPB		20 Dec 2021 09:44:42	S	Std
10.0 - 1	510	51362	-	10.00007471B	PPB		20 Dec 2021 09:46:59	S	Std
ICV18 - 1	ICV18	20242	3.9423	-7471B	PPB		20 Dec 2021 09:49:49	U	SMPL
ICB18 - 1	ICB18	9	-0.0084	-7471B	PPB		20 Dec 2021 09:54:36	U	SMPL
CCV68 - 1	CCV68	25699	5.0078	-7471B	PPB		20 Dec 2021 09:56:51	U	SMPL
CCB68 - 1	CCB68	-82	-0.0262	-7471B	PPB		20 Dec 2021 09:59:07	U	SMPL
CRA - 1	CRA	1127	0.2099	-7471B	PPB		20 Dec 2021 10:01:22	U	SMPL
HighStd - 1	HighStd	51378	10.0220	-7471B	PPB		20 Dec 2021 10:03:38	U	SMPL
ChkStd - 1	ChkStd	36237	7.0655	-7471B	PPB		20 Dec 2021 10:09:07	U	SMPL
PB141578BL - 1	PB141578BL	60	0.0015	-7471B	PPB		20 Dec 2021 10:11:22	U	SMPL
PB141578BS - 1	PB141578BS	20322	3.9579	-7471B	PPB		20 Dec 2021 10:13:40	U	SMPL
M5114-02 - 1	HS-OILY-DEBRIS	120	0.0132	-7471B	PPB		20 Dec 2021 10:15:56	U	SMPL
M5122-02 - 1	369	2554	0.4885	-7471B	PPB		20 Dec 2021 10:18:13	U	SMPL
M5124-01 - 1	F2337-STOPJOINT-TP	1955	0.3715	-7471B	PPB		20 Dec 2021 10:20:29	U	SMPL
M5142-01 - 1	A2	608211	HIGH	-7471B	PPB		20 Dec 2021 10:22:47	X	
CCV69 - 1	CCV69	25999	5.0664	-7471B	PPB		20 Dec 2021 10:26:54	U	SMPL
CCB69 - 1	CCB69	-72	-0.0243	-7471B	PPB		20 Dec 2021 10:29:09	U	SMPL
M5147-01 - 1	VNJ-210	2645	0.5063	-7471B	PPB		20 Dec 2021 10:31:25	U	SMPL
M5168-01 - 1	SB-08-C	162	0.0214	-7471B	PPB		20 Dec 2021 10:33:41	U	SMPL
M5168-01DUP - 1	SB-08-CDUP	264	0.0413	-7471B	PPB		20 Dec 2021 10:35:56	U	SMPL
M5168-01MS - 1	SB-08-CMS	19098	3.7189	-7471B	PPB		20 Dec 2021 10:38:12	U	SMPL
M5168-01MSD - 1	SB-08-CMSD	18984	3.6966	-7471B	PPB		20 Dec 2021 10:40:27	U	SMPL
M5168-02 - 1	SB-08-G	145	0.0181	-7471B	PPB		20 Dec 2021 10:42:44	U	SMPL
M5168-03 - 1	SB-21-C	7173	1.3904	-7471B	PPB		20 Dec 2021 10:45:01	U	SMPL
M5168-04 - 1	SB-21-G	5774	1.1172	-7471B	PPB		20 Dec 2021 10:47:18	U	SMPL
M5168-05 - 1	SB-01-C	318	0.0519	-7471B	PPB		20 Dec 2021 10:49:34	U	SMPL
M5168-06 - 1	SB-01-G	3509	0.6750	-7471B	PPB		20 Dec 2021 10:51:50	U	SMPL
CCV70 - 1	CCV70	26102	5.0865	-7471B	PPB		20 Dec 2021 10:54:06	U	SMPL
CCB70 - 1	CCB70	-66	-0.0231	-7471B	PPB		20 Dec 2021 10:56:22	U	SMPL
M5142-01DLX10 - 1		122310	23.8723	-7471B	PPB		20 Dec 2021 10:58:39	U	SMPL
M5168-01LX5 - 1		-688	-0.1445	-7471B	PPB		20 Dec 2021 11:00:54	U	SMPL
M5168-01A - 1		20469	3.9866	-7471B	PPB		20 Dec 2021 11:03:13	U	SMPL
M5142-01DL2X50 - 1		23215	4.5228	-7471B	PPB		20 Dec 2021 11:05:29	U	SMPL
M5146-01 - 1	NB-8-121721	639	0.1146	-7471B	PPB		20 Dec 2021 11:22:25	U	SMPL
CCV71 - 1	CCV71	25903	5.0477	-7471B	PPB		20 Dec 2021 11:24:41	U	SMPL
CCB71 - 1	CCB71	-73	-0.0245	-7471B	PPB		20 Dec 2021 11:26:58	U	SMPL