



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 12/21/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 14:15
In Date: 12/20/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:14
Out Date: 12/21/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117888

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M5160-01	EX570	1	1.14	8.80	9.94	8.74	86.4	
M5160-02	EX571	2	1.18	8.53	9.71	8.21	82.4	
M5160-03	EX572	3	1.11	8.76	9.87	8.42	83.4	
M5160-04	EX574	4	1.19	8.41	9.6	8.21	83.5	
M5160-05	EX575	5	1.14	8.64	9.78	8.4	84.0	
M5160-06	EX576	6	1.19	8.53	9.72	8.26	82.9	
M5160-07	EX577	7	1.14	8.59	9.73	8.23	82.5	
M5160-08	EX578	8	1.17	8.54	9.71	8.24	82.8	
M5160-09	EX579	9	1.1	8.79	9.89	8.12	79.9	
M5160-10	EX580	10	1.18	8.50	9.68	7.85	78.5	
M5160-11	EX581	11	1.16	8.72	9.88	8.33	82.2	
M5160-12	EX582	12	1.15	8.82	9.97	8.3	81.1	
M5160-13	EX583	13	1.17	8.81	9.98	8.74	85.9	
M5160-14	EX596	14	1.16	8.51	9.67	8.05	81.0	
M5160-15	EX596D	15	1.16	8.51	9.67	8.05	81.0	
M5160-16	EX596S	16	1.16	8.51	9.67	8.05	81.0	
M5160-17	EX597	17	1.17	8.47	9.64	7.94	79.9	
M5160-18	EX598	18	1.16	8.64	9.8	7.9	78.0	
M5160-19	EX599	19	1.15	8.41	9.56	7.86	79.8	
M5160-20	EX5A0	20	1.18	8.71	9.89	7.99	78.2	
M5160-21	EX5A1	21	1.14	8.83	9.97	8.39	82.1	
M5160-22	EX5A2	22	1.14	8.84	9.98	8.68	85.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$